



| Running Contract Details | |
|--------------------------------|--------------------------------------|
| Equipment Name | Electrical suction apparatus |
| Running Contract Valid Till | 28-11-2024 |
| Tender Ref No | KMSCL/EP/T460(R)/1273/2022 |
| Tendered Quantity | 15 |
| Supplier Name | M/s New India Surgical and Chemicals |
| GST No | 32AAGFN8445E1ZD |
| Installation & Delivery Period | 8 Week(s) |
| Up-time / PM vist | 95% & 4 Visits per year |
| Warranty period | 3 Years |

| Supplier`s Details | | |
|---|-----------------|-------------------------|
| Address | Contact Details | |
| 30/1212IV Prasanth Buildings Near Medical College Hospital Calicut-673008 | Contact Person | Prasanth N |
| | Phone | 0495-2354188 |
| | Mobile No | 9447054188 |
| | Email | nisc2000@rediffmail.com |

| Item-wise Price Details | | | | |
|-------------------------|---|---|------------------------------------|------------------|
| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
| 1 | Electrical suction apparatus <i>Model & Make : DOMINANT FLEX/ Medela AG</i> | 541000.32 Incl.GST :12% | 39898.77 | 580899.09 |
| | | 541000.32 | 39898.77 | 580899.09 |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) | | | | | | | |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Rate | 4 th Year | 5 th Year | 6 th Year | 7 th Year | 8 th Year | 9 th Year | 10 th Year |
| Electrical suction apparatus | | | | | | | |
| Labour | 16,230.00 | 17,853.00 | 19,638.00 | 21,602.00 | 23,762.00 | 26,139.00 | 28,752.00 |
| Comprehensive | 32,460.00 | 35,706.00 | 39,277.00 | 43,204.00 | 47,525.00 | 52,277.00 | 57,505.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Electrical suction apparatus

1. Should deliver high vacuum of -90Kpa/-85 Kpa, 675 mm Hg
2. Flow rate range: 35 - 60 litres/ minute and should be able to vary the range as per the requirement
3. Should provide with Piston Cylinder Technology
4. Should supply with Membrane Vacuum Regulator
5. Should be available autoclavable PSU jars of 3-5 liters (2 no's) and lids with over flow protection device
6. Should be available safety jar with over flow protection device
7. Tubing's should be made from silicone
8. Should supply with standard rail which is attached in the original company made mobile trolley.
9. Mobile trolley should have 2 Castors/ Four castors with brakes On/ Off switch.
10. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission/ FDA(US)/ STQC CB certificate/ STQC S certificate or valid detailed electrical and functional safety test report from ERTL