



Running Contract Details	
Equipment Name	Embedding Station
Running Contract Valid Till	29-09-2024
Tender Ref No	KMSCL/EP/T460/960/2022
Tendered Quantity	5
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Embedding Station <i>Model & Make : EC 500 / ESPECIALIDADES MEDICAS MYR SL</i>	1054920 Incl.GST :18%	73844.4	1128764.4
		1054920	73844.4	1128764.4

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Embedding Station							
Labour	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00
Comprehensive	30,000.00	30,000.00	30,000.00	35,000.00	35,000.00	35,000.00	40,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Embedding Station

1. Environmental operating temperature range should be +20°C to +30°C
2. Working temperature should be 50°C to 75°C (Adjustable in 1°C increments)
3. Paraffin reservoir should be minimum of 4 litres
4. Workflow should be symmetric layout
5. Cassette warming tray should be minimum of carrying 150 cassettes capacity
6. Manual & Foot Switch should be provided for dispenser
7. Display should be 5.7 inch or more capacitive, LCD touch screen
8. Timer should be programmable
9. Should have at least two collection tray for waste collection.
10. Wax reservoir filters should be available to prevent foreign bodies entering to moulds
11. Work days and time can be programmed for automatic instrument operation
12. Adjustable wax flow setting screw should be available including temperature setting for:
 - a. Left & right hand warming tray
 - b. Paraffin reservoir
 - c. Working surface
13. Removable forceps holders (for 6 forceps at a time) should be available
14. Optional position adjustable White LED light should be available
15. Continuous heated forceps plugging point 2 nos and supplied with 2 nos heated forceps of different sizes.
16. Cool spot should be provided
17. Scraper should be provided for cleaning purpose which avoids scratches Cold Plate
18. Should maintain -6°C consistent in temperature (self-regulating)
19. Work load capacity 65 blocks solidified in 30 minutes
20. Environmental operating temperature range +20°C to 30°C