

(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647 Email :engg\_ep@kmscl.kerala.gov.in CIN: U24233KL200TSGC021616, PAN : AADCK4029M, GSTIN : 32AADCK4029M1ZK

Running Contract Details				
Equipment Name	Foetal Doppler			
Running Contract Valid Till	24-11-2024			
Tender Ref No	KMSCL/EP/T455/47/2022			
Tendered Quantity	1000			
Supplier Name	M/s THIRUPATHI ENTERPRISES			
GST No	33GIQPS1238L1ZO			
Installation & Delivery Period	6 Week(s)			
Up-time / PM vist	95% & 0 Visits per year			
Warranty period	1 Years			

Supplier`s Details					
Address	Contact Details				
18/6	Contact Person	SUNDAR			
Rajakumari Towers A.K.S. Theatre Road	Phone				
Kovilpatti - 628502	Mobile No	8056453162			
	Email	thirupathisurgicals@gmail.com			

Item-wise Price Details						
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total		
1	Foetal Doppler Model & Make : CONTEC 10 CL/ CONTEC MEDICAL	3584 Incl.GST :12%	264.32	3848.32		
		3584	264.32	3848.32		

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.

4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

## **Equipment :Foetal Doppler**

- 1. Should be handy, work on rechargeable battery at least 3.5V DC
- 2. Should have an ON/OFF switch
- 3. Should have an LCD display for FHR
- 4. Should have Auto counting.
- 5. Should have probe storage facility to avoid accidental falling
- 6. Should have facility to control the volume of FHR sound
- 7. Should provide rechargeable battery along with AC adaptor for charging.
- 8. Should have Automatic power off facility.
- 9. Should have battery level indication.
- 10. Should be supplied with a carry pouch.