



Running Contract Details	
Equipment Name	OT Table Electric - B
Running Contract Valid Till	15-03-2026
Tender Ref No	KMSCL/EP/T517(R)/197B/2023
Tendered Quantity	50
Supplier Name	M/s STAAN BIO-MED ENGINEERING PRIVATE LIMITED
GST No	33AAICS8026R1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
190-A BHARATHIAR ROAD GANAPATHY COIMBATORE TAMIL NADU-641006	Contact Person	Mr M MARIA ALBERT STANLY
	Phone	0422 - 2531008
	Mobile No	9363121030
	Email	staan3@hotmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>OT Table Electric - B</b> <i>Model &amp; Make : WARRIOR /STAAN BIO- MED ENGINEERING PRIVATE LIMITED</i>	323999.68 Incl.GST :18%	22679.98	346679.66
		<b>323999.68</b>	<b>22679.98</b>	<b>346679.66</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>OT Table Electric - B</b>							
Labour	5,800.00	9,800.00	9,800.00	14,800.00	14,800.00	19,800.00	19,800.00
Comprehens ive	12,800.00	14,800.00	16,800.00	18,800.00	20,800.00	22,800.00	24,800.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :OT Table Electric - B**

1. The table should have minimum of 4 sections ie. head section, leg section, seat section and back plate section.
2. The table should be electrically operated having the following hand switch operated functions (all the dimensions will have a permitted deviation of +/- 10 %)

	Description	Range
i.	Up	1000mm Maximum
ii	Down	680mm Minimum
iii	Trendelenburg	1. deg
iv	Reverse Trendelenburg	25 deg
v	Right Lateral tilt	20 deg
vi	Left Lateral tilt	20 deg
Vii	Back up	80deg
viii	Back down	40deg

3. In addition to the above motorized operated functions, the table must have the following manual functions.
  - a. Head section tilting
  - ii. The leg section should have 90° down movement and should move side wards to a minimum of 90 degree.
4. The table should have in-built kidney bridge
5. Should have enhanced weight bearing casters fitted with ball bearing.
6. Table should have a stable braking position with single lever foot operated brake pedal.
7. The table should be supplied with the following accessories.
  - i. Mattress for the complete table top in sections - 1 set
  - ii. A pair of arm boards with pad and fixing clamp – 1
  - iii. A pair of padded shoulder support with clamps (SS grade 304) – 1
  - iv. A pair of padded lateral support with clamps (SS grade 304) – 1
  - v. A pair of padded leg crutches with clamps (SS grade 304) – 1
  - vi. Anesthetic screen frame with clamp (SS grade 304) – 1

- vii. Should supply Patient restraint strap – 1No, A pair of padded leg support to use in trendelenburg position.
- viii. The table should have a heavy and sturdy base and compact to provide adequate foot room for the operating team.
- ix. The base cover, lifting column cover and side rails should be made of stainless steel.

8. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate/ STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid

9. The table shall have a radiolucent tabletop

10. Battery backup shall be available for minimum 50 functions.

11. Should have a minimum weight bearing capacity of 200Kg

**Note:**

- 1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
- 2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail