



Running Contract Details	
Equipment Name	Gel Card Centrifuge
Running Contract Valid Till	13-02-2024
Tender Ref No	KMSCL/EP/T416(R)/1040C/2021
Tendered Quantity	20
Supplier Name	M/s Tulip Diagnostics Pvt Ltd
GST No	30AAACT1411B1ZH
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Thattil Nadakalan Complex Kuriachira P.O Thirssur - 680 006	Contact Person	Mr. D. Murali
	Phone	0487-2250793
	Mobile No	9447047079
	Email	tulipkerala@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Gel Card Centrifuge <i>Model & Make : CC 2400</i>	159300 Incl.GST :18%	11151	170451
		159300	11151	170451

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Gel Card Centrifuge							
Labour	7,000.00	7,000.00	8,000.00	8,000.00	9,000.00	9,000.00	10,000.00
Comprehensive	17,000.00	17,000.00	18,000.00	18,000.00	19,000.00	19,000.00	20,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Gel Card Centrifuge

1. Capacity
 - a. Swing out rotor should accommodate 12 or morer standard Gel cards.
2. Max speed of the centrifuge should be 1600 10 RPM
3. Max RCF should be 279 1 % g
4. Should have audio visual alarms for Imbalance, end of cycle and any other errors
5. Should have single hand rotor removal for cleaning
6. Should have LCD display with touch key pad
7. Motor should be Brushless
8. Net weight should be less than 20 kg
9. Power supply should be 220-240 volts, 50 Hz, Single phase