



Running Contract Details	
Equipment Name	X-RAY LOBBY LED VARIOUS TYPE
Running Contract Valid Till	23-04-2023
Tender Ref No	KMSCL/EP/T383/412/2020
Tendered Quantity	500
Supplier Name	M/s Bharath Business Lines
GST No	
Installation & Delivery Period	
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
2/1875A E.T.Tower Wayanad Road Kozhikode - 67300	Contact Person	Mr. Mauroof
	Phone	
	Mobile No	8289830787
	Email	bblclt2016@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>X-ray Lobby LED Double Film</b> <i>Model &amp; Make : SLIM /Natures Global Service</i>	21996.8 Incl.GST :12%	1622.26	23619.06
2	<b>X-ray Lobby LED Three Film</b> <i>Model &amp; Make : SLIM /Natures Global Service</i>	24000.49 Incl.GST :12%	1770.04	25770.53
		<b>45997.29</b>	<b>3392.3</b>	<b>49389.59</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>X-ray Lobby LED Double Film</b>				
Labour	0.00		0.00	0.00
Comprehensive	0.00		0.00	0.00
<b>X-ray Lobby LED Three Film</b>				
Labour	0.00		0.00	0.00
Comprehensive	0.00		0.00	0.00

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :X-ray Lobby LED Double Film**

#### **X-Ray Lobby(LED)-Double Film**

1. Should have double film size(14" x 17") capacities.
2. The equipment should have a high level of control luminance, without flicker, from a unit that is easy to clean and maintain.
3. The X-Ray viewing screen illumination should dimmable LED of a minimum of 60000 hours life and shall work on single-phase power supply.
4. Should have minimum 10000 Lux output adjustable.
5. Should have individual brightness & ON/OFF controls.
6. The front panel diffuser should be of a glare-free type.
7. Should have a clipless mechanism to hold & secure the X-Ray negative film when in use.
8. LED Lamps should provide a uniform level of illumination across the entire front panel diffuser and should be controlled by electronic step-less dimming controls to provide flicker-free dimming from maximum brightness to off.
9. Individual light controls for each plate.
10. Equipment shall be elegant & compact.
11. Body should be made up of mild steel powder coated/Stainless steel
12. Should be grounded properly.

### **Equipment :X-ray Lobby LED Three Film**

1. Should have three film size (14" x 17") capacities.
2. The equipment should have high level of control luminance, without flicker, from a unit that is easy to clean and maintain.
3. The X-Ray viewing screen illumination should dimmable LED of a minimum of 60000 hours life and shall work on a single-phase power supply.
4. Should have minimum 10000 Lux output adjustable.
5. Should have individual brightness & ON/OFF controls.
6. The front panel diffuser should be of a glare-free type.
7. Should have a clipless mechanism to hold & secure the X-Ray negative film when in use.
8. LED Lamps should provide a uniform level of illumination across the entire front panel diffuser and should be controlled by electronic step-less dimming controls to provide flicker-free dimming from maximum brightness to off.
9. Individual light controls for each plate.
10. Equipment shall be elegant & compact.
11. Body should be made up of mild steel powder coated/Stainless steel
12. Should be grounded properly

