



| Running Contract Details       |                                   |
|--------------------------------|-----------------------------------|
| Equipment Name                 | Laparoscope Set For Sterilization |
| Running Contract Valid Till    | 24-10-2024                        |
| Tender Ref No                  | KMSCL/EP/T442/66C/2022            |
| Tendered Quantity              | 20                                |
| Supplier Name                  | M/s Lakshmi Surgicals             |
| GST No                         | 32AACFL2603A1ZB                   |
| Installation & Delivery Period | 6 Week(s)                         |
| Up-time / PM vist              | 95% & 2 Visits per year           |
| Warranty period                | 3 Years                           |

| Supplier`s Details  |                 |                                |
|---|-----------------|--------------------------------|
| Address   | Contact Details |                                |
| 2nd Floor 12/307-C1<br>KSIE Building CIC Freight Station<br>ICTT Road Anavathil Jn.<br>Udyogamandal Eloor<br>Ernakulam - 683501 | Contact Person  | Mr. Viji Vijayan               |
|   | Phone           | 0484-2988006                   |
|   | Mobile No       | 99466 02090                    |
|   | Email           | lakshmisurgicals2005@gmail.com |

| Item-wise Price Details |                                      |   |                                    |                  |
|-------------------------|--------------------------------------|---|------------------------------------|------------------|
| #                       | Item Details                         | Unit Rate<br>(Incl.all taxes & charges) | Service Charges<br>(Through KMSCL) | Grand Total      |
| 1                       | Telescope 5mm, 30 degree             | 187054.08<br>Incl.GST :12%              | 13795.24                           | 200849.32        |
| 2                       | Ring Applicator 6 to 7 mm            | 38642.58<br>Incl.GST :12%               | 2849.89                            | 41492.47         |
| 3                       | Trocar & Cannula 7mm                 | 31362.8<br>Incl.GST :12%                | 2313.01                            | 33675.81         |
| 4                       | Trocar & Cannula 5 to 6mm            | 31362.8<br>Incl.GST :12%                | 2313.01                            | 33675.81         |
| 5                       | Verees pneumoperitoneum needle 8 cm  | 9660<br>Incl.GST :12%                   | 712.43                             | 10372.43         |
| 6                       | Verees pneumoperitoneum needle 10 cm | 9660<br>Incl.GST :12%                   | 712.43                             | 10372.43         |
| 7                       | Verees pneumoperitoneum needle 12 cm | 11592<br>Incl.GST :12%                  | 854.91                             | 12446.91         |
|                         |                                      | <b>319334.26</b>                        | <b>23550.9</b>                     | <b>342885.16</b> |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) |                      |                      |                      |                      |                      |                      |                       |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Rate   | 4 <sup>th</sup> Year | 5 <sup>th</sup> Year | 6 <sup>th</sup> Year | 7 <sup>th</sup> Year | 8 <sup>th</sup> Year | 9 <sup>th</sup> Year | 10 <sup>th</sup> Year |
| <b>Telescope 5mm, 30 degree</b>                      |                      |                      |                      |                      |                      |                      |                       |
| Labour   | 2,852.00             | 2,852.00             | 2,852.00             | 2,852.00             | 2,852.00             | 2,852.00             | 2,852.00              |
| Comprehensive  | 8,554.00             | 8,554.00             | 8,554.00             | 8,554.00             | 8,554.00             | 8,554.00             | 8,554.00              |

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Telescope 5mm, 30 degree**

I. The following items to be supplied as one kit for lap sterilization

1. Telescope 30 degree 5mm

a. Should be a rigid telescope with 5mm diameter

b. Should have a minimum working length of 30 cm.

c. Should be supplied with 1 telescope with viewing direction of and 30°.

d. Should have incorporated fiber optic light transmission.

e. Should have scratch resistant tip.

f. Should be autoclavable

2. Ring Applicator 6 to 7mm

a. Ring applicator, may be fitted with 2 silastic rings. For use with trocar size 6mm

3. Trocar & Cannula

a. Trocar, Size 7 mm consisting of : Trocar only, with pyramidal tip, cannula without valve, with insufflations stop cock, length 8 cm or more , Multifunctional Valve - 4 Nos

b. Trocar, Size 5 to 6mm consisting of : Trocar only, with pyramidal tip, cannula without valve, with insufflations stop cock, length 8

cm or more , Multifunctional Valve - 4 Nos

4. Verees pneumoperitoneum needle – Small, medium and Large

a. Verees pneumoperitoneum needle with spring action blunt stylet, length should be atleast 8 cm, 10 cm and 12 cm, , should have a luer -lock connector for insufflations.

II. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US). Copy of the certificate/ test report shall be produced along with the technical bid