



Running Contract Details	
Equipment Name	Semi Fowlers Cot
Running Contract Valid Till	27-02-2026
Tender Ref No	KMSCL/EP/T512/1190B/2023
Tendered Quantity	1000
Supplier Name	M/s HOSPITAL BEDS
GST No	32AAXPF8778M1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
NALLALAM MINI INDUSTRIAL ESTATE MODERN INDUSTRIAL ESTATE NALLAM CALICUT-673027	Contact Person	NALINAKSHAN
	Phone	
	Mobile No	9495863827,4952422335
	Email	hospital.beds@hotmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Semi Fowlers Cot</b> <i>Model &amp; Make : HOSPITAL BEDS</i>	19800.4 Incl.GST :18%	1386.03	21186.43
		<b>19800.4</b>	<b>1386.03</b>	<b>21186.43</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year
<b>Semi Fowlers Cot</b>			
Labour	250.00	500.00	750.00
Comprehensi ve	500.00	1,000.00	1,500.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Semi Fowlers Cot**

#### **I. Semi-Fowlers Cot**

1. Should have overall approximate size of 198 x 90 x 60 cms. Size.
2. The back section should be adjustable to at least 60 degrees.
3. Top sheet shall be made of perforated 18 gauge CR sheet.
4. Should have 30mm x 50mm x 16 gauge bed frame with two cross bar support.
5. Legs shall be of 30mm x 16 gauge round CR pipe, with 30mm x 16 gauge cross bar round tube support for leg support and detachable and collapsible side rail made of SS or Aluminium.
6. Should have Hook on leg side both ends for hanging uro bags.
7. Should be provided with high quality suitable mattress of at least 4 inches thick PU foam of 40 density covered with soft water proof material, bacteriostatic & pillow.
8. All MS components should be 7 dip tank pre-treated and epoxy powder coated with 50-60 microns.
9. Should have provision for fixing IV stand on all four corners. Should be provided with one telescopic IV stand with four hook.
10. Should be mounted on four swiveling PU castors of diameter 10 to 12.5 cms.
11. Should be provided with SS laminated head and foot board.

#### **Note:**

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail