



Running Contract Details	
Equipment Name	Pacemaker ( Temporary ) - Single Chamber
Running Contract Valid Till	22-08-2024
Tender Ref No	KMSCL/EP/T416/303A/2021
Tendered Quantity	50
Supplier Name	M/s India Medtronic Pvt Ltd
GST No	27AAACI4227Q1Z8
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
1241 Solitaire Corporate Park Building No.12 4th Floor Andheri-Ghatkopar Link Road Andheri - East Mumbai - 400093	Contact Person	Pradeep Kumar S G
	Phone	02233074700
	Mobile No	9884966277
	Email	s.g.pradeep.kumar@medtronic.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Pacemaker ( Temporary ) - Single Chamber</b> <i>Model &amp; Make : 53401 / Medtronic</i>	105000 Incl.GST :5%	8260	113260
2	<b>Patient cable</b> <i>Model &amp; Make : 5492VL</i>	8850 Incl.GST :18%	619.5	9469.5
		<b>113850</b>	<b>8879.5</b>	<b>122729.5</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Pacemaker ( Temporary ) - Single Chamber</b>							
Labour	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Comprehensive	2,950.00	3,097.00	3,251.00	3,413.00	3,583.00	3,762.00	3,950.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Pacemaker ( Temporary ) - Single Chamber**

1. Should be a Single Chamber Pacemaker (Temporary) for bradycardia treatment before, during or after a surgery.
2. Stimulation burst and permanent stimulation should be available for high pacing rate.
3. Should be compact & easy-to-operate device, particularly suitable for emergency treatments.
4. Safety features, including automatic lead and battery check.
5. Should have continuous monitoring of the battery voltage.
6. Should have shock and water-resistant housing.
7. Should have back up pacing during battery change.
8. Should have Modes AOO, AAI, VOO, VII
9. Should have pacing rate 40-180 ppm.
10. Should have fast pacing (Burst rate) of 80-200 ppm.
11. Should have Pulse Amplitude of 0.1-17V
12. Should have sensitivity 1.0-20mV
13. Should have minimum battery backup > 200 hours.
14. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
15. Temporary pacing wires- 2 nos should be supplied along with each unit.