



Running Contract Details	
Equipment Name	Flexible Urethroscope
Running Contract Valid Till	28-03-2025
Tender Ref No	KMSCL/EP/T464/270A/2022
Tendered Quantity	20
Supplier Name	M/s Lakshmi Surgicals
GST No	32AACFL2603A1ZB
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
2nd Floor 12/307-C1 KSIE Building CIC Freight Station ICTT Road Anavathil Jn. Udyogamandal Eloor Ernakulam - 683501	Contact Person	Mr. Viji Vijayan
	Phone	0484-2988006
	Mobile No	99466 02090
	Email	lakshmisurgicals2005@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Flexible Urethroscope</b> <i>Model &amp; Make : Karl Storz Flexible Ureteroscope/ Karl Storz GmbH &amp; Co</i>	739200 Incl.GST :12%	54516	793716
		<b>739200</b>	<b>54516</b>	<b>793716</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Flexible Urethroscope</b>							
Labour	18,000.00	18,000.00	18,000.00	18,000.00	18,000.00	18,000.00	18,000.00
Comprehensive	60,000.00	66,000.00	72,600.00	79,860.00	87,846.00	96,631.00	1,06,294.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Flexible Urethroscopy**

#### **Equipment:**Flexible Ureteroscopy

1. Should be a flexible fiberoptic Uretero – Renoscope with Continuous Irrigation.
2. Outer diameter of the shaft should be between 6 Fr. - 10 Fr.
3. Should have an instrument channel between 3 Fr. to 4 Fr.
4. Should have 680mm  $\pm$ 20mm working length..
5. Should have a ceramic liner in the distal end of the working channel to protect it from thermal or electro Cautery damage.
6. Should have angle of view between 80- 90 field of view 00 with eyepiece adjustment.
7. Should have minimum angle of deflection 180 degree up, 270 degree down.
8. Should have active deflection mechanism.
9. Should be supplied with gas sterilization valve.
10. Should be waterproof and fully immersible in solution.
11. Should be adhering to sterilization method with ETO, FO gas.
12. Should be supplied with the following accessories of compatible sizes.
  - a. Grasping Forceps
  - b. Biopsy Forceps
  - c. Pressure compensation cap
  - d. Leakage Tester
  - e. Cleaning Brush
  - f. Carry Case
  - g. Access sheath
13. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid

detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid