



Running Contract Details	
Equipment Name	Direct Ophthalmoscope
Running Contract Valid Till	20-06-2025
Tender Ref No	KMSCL/EP/T481/35/2022
Tendered Quantity	500
Supplier Name	M/s Biowel Inc
GST No	32AAHFB7582RIZT
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Sibi Bhavan Plamoodu Pattom (PO) Trivandrum-695004	Contact Person	Unni
	Phone	0471-2304055
	Mobile No	9495931992, 9188724992
	Email	sales@biowel.com, mkt@biowel.com, biowelinc@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Direct Ophthalmoscope</b> <i>Model &amp; Make : Keeler Practitioner Ophthalmoscope (Bulb)/(LED) Part No: 1127-P 1009/1127-P1006 &amp; 1941-P-1341 / Keeler Ltd UK</i>	20158.88 Incl.GST :12%	1486.72	21645.6
		<b>20158.88</b>	<b>1486.72</b>	<b>21645.6</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>Direct Ophthalmoscope</b>				
Labour	1,000.00	1,000.00	1,000.00	1,000.00
Comprehens ive	1,500.00	1,500.00	1,500.00	1,500.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Direct Ophthalmoscope**

1. Should be rechargeable battery with Charger.
2. Should have halogen / LED light source
3. Should have red-free filters
4. Should have small and large spot sizes, fixation targets, slit aperture, and cobalt blue filter.
5. Should have wheel control with lens powers ranging from -25D to +40D in single dioptr steps up to 10D and 5D steps above that.
6. Should have illuminated lens dial.
7. Should have rubber brow rest.
8. Should have dust free optics and a spherical optical system.
9. Should be supplied with a carrying case.
10. If halogen lamp is used, then the following additional accessories should be supplied
11. Bulb – 2 nos