



Running Contract Details	
Equipment Name	Labour suite
Running Contract Valid Till	18-10-2025
Tender Ref No	KMSCL/EP/T494(R)/967/2023
Tendered Quantity	100
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Labour suite <i>Model & Make : PD-200/Paramount Bed India Ltd</i>	336300 Incl.GST :18%	23541	359841			
		336300	23541	359841			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Labour suite							
Labour	5,700.00	5,700.00	5,700.00	5,700.00	5,700.00	5,700.00	5,700.00
Comprehensi ve	14,250.00	14,250.00	14,250.00	14,250.00	14,250.00	14,250.00	14,250.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Labour suite

1. It should have three sections and sectional mattresses. The seat section should have large perineal cut. Should have 3 inches thick washable and detachable mattress.
2. Should have control device for back adjustments through remote control
3. Should be able to give trendelenburg, reverse trendelenburg and 65 degree sitting
4. Should be able to hold minimum of 200Kg.
5. It should have foldable ABS side support.
6. It should have head board (ABS) which can be detached, padded.
7. Should have wheels provided with locking system
8. Should have retractable foot section so as to convert bed into table
9. Should have infusion rods which have adjustable heights, quick release and attaches to both sides of bed
10. Leg section(foot support) should be foldable or telescopic under the mainframe completely for lithotomy positions. To and fro motion of the leg section should be very smooth.
11. Should have push grip handles
12. Should have stainless steel bowl at perineal part of the table
13. Should have catheter bag holder which can be attached on either side of bed.
14. Should have adjustable foot supports for nursing staff
15. Should be easy to clean, sterilize (especially blood stains) and maintain
16. Frame should be of epoxy powder coated steel
17. Should have dimensions of at least length 6 feet and width 2.5feet.
18. Should have power input 220 – 240 V AC, 50HZ fitted with Indian plug
19. Should have UPS of suitable rating with voltage regulation and spike protection for 60 minutes backup.
20. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
21. All consumables required for installation and standardization of the system should be included in the offer. Should be provided with a pair of leg crutches, IV pole with 4 hooks