



Running Contract Details	
Equipment Name	CPM Lower Limb (Knee Joint)
Running Contract Valid Till	05-01-2025
Tender Ref No	KMSCL/EP/T463/24A/2022
Tendered Quantity	25
Supplier Name	M/s Technomed Electronics
GST No	
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Plot No.41 Functional Industrial Estate Perungudi CHennai - 600096 Tamilnadu	Contact Person	
	Phone	
	Mobile No	
	Email	sales@technomedelectronics.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	CPM Lower Limb (Knee Joint) <i>Model & Make : TECH CPM/ Technomed</i>	59920 Incl.GST :12%	4419.1	64339.1
		59920	4419.1	64339.1
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
CPM Lower Limb (Knee Joint)				
Labour	1,620.00	1,620.00	1,620.00	
Comprehens ive	3,240.00	3,240.00	3,240.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :CPM Lower Limb (Knee Joint)

1. Unit should be leg adjustable
2. Clause deleted.
3. Should be microprocessor control having digital display for treatment time flexion angle, extension angle and speed.
4. Should be easily programmable
5. Should have an emergency 'stop' patient switch.
6. The leg rest should be adjusted simply and quickly.
7. Speed of the leg set should be adjustable by speed control at the front panel.
8. Should have a treatment time ranging from 15 to 60 minutes.
9. Should have minimum of 0 to 110 degree extension angle.
10. Leg guides should have suspension system to avoid pressure mark
11. Unit should have movement protector and spasm controlled
12. Should be operated on 50Hz 230V single phase supply.
13. Should have safety certificate from a competent authority.