



Running Contract Details	
Equipment Name	Plaster cutter
Running Contract Valid Till	31-07-2025
Tender Ref No	KMSCL/EP/T489(R)/103/2023
Tendered Quantity	30
Supplier Name	M/s UK Surgicals
GST No	32AADFU2199K1ZX
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Perumby Nalukettu Koratty Thrissur	Contact Person	Anilkumar Uk
	Phone	
	Mobile No	9946148248
	Email	sales@uksurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Plaster cutter</b> <i>Model &amp; Make : UKS1031/UK Surgicals</i>	12000.24 Incl.GST :12%	885.02	12885.26
		<b>12000.24</b>	<b>885.02</b>	<b>12885.26</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Plaster cutter**

#### **Equipment: Plaster Cutting Machine**

1. Should have a fiber body.
2. Should be able to cut fiber gauzes.
3. Blades should be corrosion resistant and highly durable.
4. Blades should have hexagonal mounting hole.
5. Should be supplied with 84mm, 74mm, and 64mm diameter blades.
6. Should be supplied with required tools for replacing the blades, brush and duster.
7. Should be supplied with carrying case to accommodate the plaster cutter and other accessories.
8. Should have a protective guard
9. Should be oscillating type
10. Should work with input 200 to 240Vac 50 Hz supply.
11. Additional Accessories
  - a. Blades (84mm) - 6Nos

**Note:** If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid