



Running Contract Details	
Equipment Name	RF Excited CO2 Laser
Running Contract Valid Till	29-08-2026
Tender Ref No	KMSCL/EP/T539/1049/2024
Tendered Quantity	5
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	<b>RF Excited CO2 Laser</b> <i>Model &amp; Make : Futura RF 50 / Derma India</i>	1170400 Incl.GST :12%	86317	1256717			
		<b>1170400</b>	<b>86317</b>	<b>1256717</b>			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>RF Excited CO2 Laser</b>							
Labour	31,350.00	31,350.00	31,350.00	31,350.00	31,350.00	31,350.00	31,350.00
Comprehensive	83,600.00	1,04,500.00	1,25,400.00	1,25,400.00	1,25,400.00	1,25,400.00	1,25,400.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :RF Excited CO2 Laser**

#### **Equipment: RF Excited CO2 Laser**

1. Laser power output up to power 50W
2. Power Supply AC 220V 50Hz <+5%
3. Beam mode TEM00. Beam diameter 3.5mm. Modulating frequency 0-20 KHz.
4. Spot size 0.1mm/1.25mm (adjustable)
5. Spot patter circular/triangle/square/rectangle/rhombus/line.
6. Spot density X1 to X6. Aiming beam 635nm, ?5mW.
7. LASER operating mode normal/ultra pulse/fractional.
8. Scanning mode, randomized scanning.
9. Pulse interval 1W~50W (every stepping: 1W).Pulse energy Max 50W.