

Running Contract Details	
Equipment Name	Ophthalmic Operating microscope Model B
Running Contract Valid Till	19-01-2026
Tender Ref No	KMSCL/EP/T505(R)/87B/2023
Tendered Quantity	10
Supplier Name	M/s Serumbiologicals
GST No	32FKIPS9153R2Z4
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Rajagiri Southgat Thirunakkara Kottayam	Contact Person	Sunilkumar
	Phone	4812563636
	Mobile No	9745163636
	Email	serumbiologicals@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Ophthalmic Operating microscope Model B <i>Model & Make : OPMI LUMERA 300 BASIC SYSTEM WITH XY UNIT/Carl Zeiss</i>	2330000.01 Incl.GST :18%	163100	2493100.01
		2330000.01	163100	2493100.01

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Ophthalmic Operating microscope Model B							
Labour	81,550.00	81,550.00	87,375.00	93,200.00	99,025.00	1,04,850.00	1,10,675.00
Comprehensive	1,74,750.00	1,86,400.00	1,98,050.00	2,09,700.00	2,15,525.00	2,21,350.00	2,27,175.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Ophthalmic Operating microscope Model B

1. Main microscope – Apochromatic optics – manual 5 step magnification changer.
2. Focusing – motorized focusing via foot control- range 40 mm.
3. XY coupling – adjustment range 40 mm x 40 mm or more – Button for starting positions of X-Y coupling.
4. Illumination – LED illumination with filters for halogen like illumination, yellow filter (blue light filtering) and color enhancement filter- variable red reflex enhancement- LED light source should have auto backup option.
5. Objectives Lens-f= 200 mm, should have a diameter of 65 mm.
6. Binocular tube- 45 deg inclined binocular tube.
7. Eye pieces – 10 x/12.5 x wide field eye pieces.
8. Stand- floor stand with 4 large wheels – overall stand height 1.7 meter or more and arm length 1.0 meter or more
9. Lamp should be switched off when suspension arm is raised to full height
10. Safety features- Microscope arm should have a limit of downward movement lock as to prevent the microscope getting into contact with the patient even if it is inadvertently lowered.
11. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate /STQC S certificate or valid detailed electrical and functional safety test report from ERTL.

Note:

1. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail
2. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid