



Running Contract Details	
Equipment Name	Shock Wave Therapy Unit
Running Contract Valid Till	13-03-2025
Tender Ref No	KMSCL/EP/T475/1626/2022
Tendered Quantity	10
Supplier Name	M/s Innovation Healthcare
GST No	33AAFFI7509B1ZS
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
5277 First floor Ram nagar north extension 8th street Madipakkam Chennai-600091	Contact Person	Manimaran
	Phone	
	Mobile No	9047489400
	Email	Innovationhealthcare2014@gmail.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	<b>Shock Wave Therapy Unit</b> <i>Model &amp; Make : OCE_ESWT_2/Oceanus</i>	446500 Incl.GST :12%	32929.37	479429.37			
		<b>446500</b>	<b>32929.37</b>	<b>479429.37</b>			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
Shock Wave Therapy Unit							
Labour	2,542.37	2,542.37	2,542.37	2,542.37	2,542.37	2,542.37	2,542.37
Comprehensive	21,186.44	21,186.44	21,186.44	25,423.73	25,423.73	25,423.73	29,661.02

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Shock Wave Therapy Unit**

1. Standard unit delivering acoustic shock waves
2. Max. pressure up to : > 4 bars
3. Frequency 1 – 15 or more
4. Different Modes : Single mode & Continuous mode, Burst mode, Intensity Gradient mode / Equivalent
5. colour touch screen > 5.5”
6. Preset programs based on indications
7. Mains supply 220 - 230 V, 50–60 Hz
8. Standard accessories
9. Should have safety certificate from a competent authority, CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/  
test report shall be produced along with the technical bid