



Running Contract Details	
Equipment Name	Ophthalmic Operating Microscope
Running Contract Valid Till	03-06-2026
Tender Ref No	KMSCL/EP/T529(R)/87/2024
Tendered Quantity	25
Supplier Name	M/s Appasamy Associates PVT. LTD
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Ophthalmic Operating Microscope <i>Model & Make : AAOM Brilliant Advent FS 9/Appasamy</i>	879200 Incl.GST :12%	64841	944041
		879200	64841	944041

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Ophthalmic Operating Microscope							
Labour	31,400.00	39,250.00	47,100.00	54,950.00	62,800.00	70,650.00	78,500.00
Comprehensive	62,800.00	70,700.00	78,500.00	86,400.00	94,200.00	1,02,100.00	1,09,900.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Ophthalmic Operating Microscope

1. Should have apochromatic and imported optics
2. Should have 5 step magnification
3. Should have working distance of objective lens of focal length $F = 200\text{mm}$
4. Eye piece should be minimum 10x ,12.5x or 16x wide
5. Should have XY coupling
6. Should have tiltable binocular with converging optics.
7. Should have field of view from at least 15mm to 50mm
8. Should have cold light/LED coaxial illumination
9. Should have tools free design for stand-by bulb change over and for fai bulb replacement if required.
10. Should have heat absorbing and UV filters.
11. Should be floor standing type with fiber wheels with brake.
12. Should have a minimum vertical stroke of 400mm
13. Should have rust free stand.
14. Should be operated in 200-240 Vac 50/60 Hz input supply.
15. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail