



Running Contract Details	
Equipment Name	Retinal Camera for ROP Screening
Running Contract Valid Till	29-09-2024
Tender Ref No	KMSCL/EP/T446/168A/2022
Tendered Quantity	15
Supplier Name	M/s Forus Helath Pvt Ltd
GST No	29AABCF4661J1ZE
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
#2234 23rd Cross Banashankri 2nd Stage Next to Meenakshi ENT center Karnataka.	Contact Person	KMK Rajendran
	Phone	08041624041
	Mobile No	9741026644
	Email	kmk@forushealth.com, azmath@forushealth.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Retinal Camera for ROP Screening</b> <i>Model &amp; Make : 3 nethra neo / Forus Health Private Limited</i>	2004800 Incl.GST :12%	147854	2152654
		<b>2004800</b>	<b>147854</b>	<b>2152654</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Retinal Camera for ROP Screening</b>							
Labour	95,580.00	1,11,510.00	1,27,440.00	1,43,370.00	1,59,300.00	1,75,230.00	1,91,160.00
Comprehen sive	1,27,440.00	1,43,370.00	1,59,300.00	1,91,160.00	2,23,020.00	2,54,880.00	2,86,740.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Retinal Camera for ROP Screening**

1. Should have a portable type with carrying case.
2. Screening age should be lesser than 55 weeks of PMA
3. Minimum pupil diameter 4.0 mm
4. Should have minimum 4 Megapixel or better camera.
5. Camera resolution should have minimum 2040 x 2040 – 24 bitcolor
6. Light source: LED
7. Should have warm white colour with light guide
8. Should have minimum, 8000 lux at contact lens
9. Should have cone angle 98 degree and cone diameter 33 mm or better.
10. Contact lens diameter should have 10 mm and imaging surface radius should be 6 mm.
11. Field of view should have 120°(from the centre of the eye )or higher degree
12. Should have singlehanded operation.
13. Should have facility to imaging stills and videos.
14. Should have focusing control through foot pedal
15. Light intensity can be controlled with foot pedal
16. Probe weight should be less than 350 gm.
17. Should provide suitable printer with equipment.
18. Power supply AC 100 – 240V, 50/60 Hz
19. Should have safety certificate from a competent authority, CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
20. The supplier should arrange training for doctors in a reputed institutions anywhere in India free of cost.

