



Running Contract Details	
Equipment Name	Deep Freezer (-20C)
Running Contract Valid Till	24-11-2024
Tender Ref No	KMSCL/EP/T455/170/2022
Tendered Quantity	50
Supplier Name	M/s New India Surgicals and Chemicals
GST No	32AAGFN8445E1ZD
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
30/1212IV Prasanth Buildings Near Medical College Hospital Calicut-673008	Contact Person	Prasanth N
	Phone	0495-2354188
	Mobile No	9447054188
	Email	nisc2000@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Deep Freezer (-20C) <i>Model & Make : DW-YL27OH / Blue Star</i>	150563.28 Incl.GST :18%	10539.43	161102.71
		150563.28	10539.43	161102.71
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
	Deep Freezer (-20C)			
Labour	4,000.00	5,000.00	5,250.00	
Comprehensive	10,000.00	10,000.00	10,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Deep Freezer (-20C)

1. Capacity : 200 litre
2. Inner shelf : 3 Nos
3. Refrigerant : CFC free
4. Temperature control: Micro-processor controlled digital display with temperature resolution of 0.1 degree C
5. Alarms: high/ low temperature
6. Door closing and locking with key
7. Adjustment: adjustable leveling feet standard (optional casters)
8. Power supply: 210-240V/50-60Hz.
9. Should be a vertical type freezer
10. Inner chamber should be made of SS and outer should be GI/CRCA powder coated
11. The product should be CE Certified.