



Running Contract Details	
Equipment Name	LaborCot
Running Contract Valid Till	24-11-2024
Tender Ref No	KMSCL/EP/T455/64D/2022
Tendered Quantity	25
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	LaborCot <i>Model & Make : Medik Enterprises</i>	32568 Incl.GST :18%	2279.76	34847.76
		32568	2279.76	34847.76

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
LaborCot				
Labour	300.00	300.00	300.00	300.00
Comprehensive	850.00	850.00	850.00	850.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :LaborCot

1. Should have a SS table top with all edges smoothly finished
2. Should have a dimension of L 180 X W 70 X H 80 cm
3. Should have a drainage hole of diameter 12cm-basin type and collecting tray of stainless steel
4. The U cut should have a diameter of 35 cm.
5. Should provide with a detachable saline stand having a provision for fixing four positions (Head end and middle position) on the cot.
6. Base should be mounted on four heavy duty rubber shoes.
7. Should be as per ISI standard.
8. Mattress (40mm thickness) on the entire top of Labor Cot
9. Should provide a pair of padded Lithotomy leg clutches with fixing clamps and a pair of hand grips
10. Back rest shall be adjustable upwards

Note: The Stainless steel shall be of grade SS304. Material test certificate for the proof of usage of SS304 shall be furnished along with invoice and the manufacturer shall also submit undertaking that the tested material is used for the construction of the equipments supplied