



Running Contract Details	
Equipment Name	HOT AIR OVEN (LAB)-VARIOUS TYPES
Running Contract Valid Till	03-10-2024
Tender Ref No	KMSCL/EP/T400/56/2021
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details					
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Hot air oven (Lab) <i>Model & Make : MHOS -4 & MHOS -3/Labline</i>	200	25370 <small>Incl.GST :18%</small>	1775.9	27145.9
2	Hot air oven (Lab) Small <i>Model & Make : MHOS -4 & MHOS -3/Labline</i>	200	32332 <small>Incl.GST :18%</small>	2263.24	34595.24
			57702	4039.14	61741.14

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Hot air oven (Lab)				
Labour	3,000.00		3,000.00	3,000.00
Comprehensive	5,000.00		6,500.00	7,500.00
Hot air oven (Lab) Small				
Labour	3,000.00		3,000.00	3,000.00
Comprehensive	5,000.00		6,500.00	7,500.00

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
ve			

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Hot air oven (Lab)

. Hot Air Oven (450x450x450mm)

1. Should be operated on 230V, 50Hz single-phase AC supply and having temperature ranging between 50-200°C
2. Should be made of the double-walled chamber -Inner made of stainless steel SS 304 grade and powder-coated outer surface.
3. Should provide three heating elements on three sides of the equipment for uniform temperature on all shelves.
4. Should be provided with air circulating fan.
5. Should provide a variable microprocessor-based digital temperature controller with a digital display and thermometer should be provided separately.
6. Should have a minimum chamber size of (L*B*H) 450*450*450 with 2 stainless steel trays with holes.
7. Should provide with air ventilation.
8. Should have a safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid

Note : L1 shall be declared separately for each item.

Equipment :Hot air oven (Lab) Small

I. Hot Air Oven(450x450x600mm)

1. Should be operated on 230V, 50 Hz single-phase AC supply and having temperature ranging between 50-200°C.
2. Should be made of a double chamber Inner made of 304-grade Stainless Steel and powder-coated outer surface.
3. Should provide with heating elements on three sides of the chamber for uniform temperature on all shelves.

4. Should be provided with air circulating fan.
5. Should provide with a variable microprocessor-based digital temperature controller with a digital display.
6. Should have a minimum chamber size of (L x B x H) 450x450x600mm with 3nos. 304-grade stainless steel trays with holes.
7. Insulating material should be thick glass wool with a silicone rubber door gasket.
8. Should be provided with air ventilation.
9. Should have provision for monitoring the temperature with a thermometer
10. Should have safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted.
Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid