



Running Contract Details	
Equipment Name	Patient warmer
Running Contract Valid Till	06-05-2023
Tender Ref No	KMSCL/EP/T393/165/2021
Tendered Quantity	1000
Supplier Name	M/s Fortuna Technologies
GST No	32AABFF0816G1Z3
Installation & Delivery Period	
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
PTPA-21 - Pipeline Road PTP Avenue Trivandrum-695030 Kerala	Contact Person	C R Nair
	Phone	2364573
	Mobile No	9847944991
	Email	fortuna@asianetindia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Patient warmer <i>Model & Make : Nellcor WT6000/Medtronic</i>	78960 Incl.GST :12%	5823.3	84783.3
2	Cost of additional Blanket disposable adult full body	448 Incl.GST :12%	33.04	481.04
3	Cost of additional Blanket disposable adult Upper half	649.6 Incl.GST :12%	47.91	697.51
		80057.6	5904.25	85961.85

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
Patient warmer			
Labour	4,200.00	4,500.00	4,750.00
Comprehensi	5,500.00	5,750.00	6,000.00

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
ve			

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Patient warmer

1. Microprocessor Based.
 2. Unique Temperature Control System.
 3. High Air Flow.Temperature range: 34- 42°C
 4. Alarm indicator for Over Temperature and Under Temperature.
 5. Temperature Digital Display.
 6. In-built Air Filter.
 7. Elapsed Time Display.
 8. IV pole/bedrail is mountable.
 9. Accessories : 3 Nos. of blanket disposable adult full body
: 3 Nos. of the blanket disposable upper half
1. Should work on 220-240V AC/50Hz.
 2. Should have a safety certificate from a competent authority CE issued by a notified body registered in the European commission/FDA(US)/STQC CB certificate /STQC S certificate or a valid detailed electrical and functional safety test report from ERTL. Copy of the certificate /test report shall be produced along with the technical bid