



Running Contract Details	
Equipment Name	MULTIFUNCTIONAL COT
Running Contract Valid Till	30-10-2025
Tender Ref No	KMSCL/EP/T493/58D/2023
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s Midmark (India) Private Limited
GST No	24AACCM0219N1ZM
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	5 Years

Supplier`s Details		
Address	Contact Details	
A 1 3 GIDC UMBERGAON Valsad Gujarat - 396171	Contact Person	Mr. Renjith Lal
	Phone	+91 260 2562 406
	Mobile No	9605680008
	Email	pkadam@midmark.com

Item-wise Price Details					
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Multifunctional Cot <i>Model & Make : Model No. DX4311 XR Semi motorized bed/M/s. Midmark (India) Private Limited</i>	100	84522.22 Incl.GST :18%	5916.56	90438.78
2	Bedside Locker Type B <i>Model & Make : AX1009B/Deluxe Bedside Locker with lock</i>	100	7743.16 Incl.GST :18%	542.02	8285.18
			92265.38	6458.58	98723.96

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year	11 th Year	12 th Year
Multifunctional Cot							
Labour	1,074.00	1,074.00	1,074.00	1,074.00	1,074.00	1,074.00	1,074.00
Comprehens ive	1,433.00	1,433.00	1,433.00	1,433.00	1,433.00	1,433.00	1,433.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment : Multifunctional Cot

MULTI-FUNCTIONAL COT	
Technical Specifications : Multifunctional couch	
	Please quote the model which is including the latest state of art equipment and meets the basic minimum specifications. The quoted model should be latest brand new.
1	It should be a semi motorized bed with motorized backrest up-down, height adjustment, knee rest up-down.
2	Should have manual adjustments for Trendelenburg and reverse Trendelenburg
3	Should have minimum 4 section lying surface which is easily removable, washable to maintain hygiene
4	Should have bumpers at each corner with excellent shock absorbing property
5	Should have 4 castor wheels to provide smooth and optimal movement, two with brakes & two without brakes
6	The back rest should move up to 70° , knee rest should move up to 24°
7	Bed should have side board with drop down mechanism with a safety feature of preventing patient falling off the bed full length bed side railing collapsible
8	Should have CPR lever to make quick release of the back rest
9	Both the head & leg board should be removable without locking mechanism for ease of use during emergency.
10	Bed should be equipped with a strong & sturdy IV pole to withstand maximum weight if infusion pumps are attached
11	Should have provision for accessories like urine bag holder, ICD Bag etc.
12	All electrical parts should have liquid ingress protection as per IP4.
13	Should work on mains 230VAC/ 50-60Hz
14	Maximum patient load- 130Kgs
15	Body should be powder coated with antimicrobial and epoxy polyester to control bacterial growth
16	Bed should be provided with, remote control, IV pole, battery backup, oxygen cylinder holder, document holder, X Ray cassette holder.
17	Bed should be compliant with ISO 13485, IEC 60601-2-52 standards and with current protection level of class 1 and shock protection level of class B
18	PU foam mattress with 40 densities should be provided.
19	Zip & stitches for the mattress cover should be concealed.
20	Should be USFDA approved/ CE marked.
21	Please specify the dimensions

	Accessories:
	Remote control - 1 no per couch
	Should perform calibration of the equipment yearly during warranty & free service period. Testing & measuring equipment used should be traceable to SI units through National /international Standards, if required
	Please specify list of consumables/ consumable spares (i.e. spares need to be replaced at regular intervals, may be quarterly/half yearly/ yearly such as annual maintenance kit etc.) if any.
	Please specify footprint size & its weight.
	Demo of the quoted model will be mandatory at the cost of the bidder if so desired by the user, after the opening of technical bid and prior to the opening of financial bid. This is for technical evaluation.
	Back to back assurance to be taken by the supplier from OEM to supply spares for minimum 12 years and to be submitted

Equipment :Bedside Locker Type B

Equipment: Bed Side Locker Type B

1. Overall dimensions should be minimum 400 W x 400 D x 800mm H
2. The Locker should have one drawer unit and one cabinet unit fitted with Cam lock.
3. The locker top should be of ABS/ LDPE sheet and the sheet should be finished in buff-matt finish.
4. The top should have a raised border of 10mm height on three sides minimum.
5. Drawer should be made 1 mm thick CRCA sheet and it should be fitted with ball slides for smooth drawer movement and also a recess to serve as handle.
6. Cabinet should made 1 mm thick CRCA sheet, fitted with hinge door and lock and should have louvers for ventilation.
7. Back of the locker should be provided with vertical spacer made of nylon with uniformly distributed total load bearing capacity of 20 kg.
8. Test certificates of these tests should be available proof loading test, cycle tests, impact test, salt spray test, castor break test.
9. Should have a cloth hanger