



**Running Contract Details**

Equipment Name	Bronchoscope Cryosurgery Unit
Running Contract Valid Till	21-12-2024
Tender Ref No	KMSCL/EP/T451(R)/1035/2022
Tendered Quantity	5
Supplier Name	M/s Aswini Health Specialities India Pvt Ltd
GST No	32AAGCA4309K1ZW
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

**Supplier`s Details**

Address	Contact Details	
Door No.IV/675 Ferry Road Kalamassery Ernakulam	Contact Person	Manoj K A
	Phone	04842550121
	Mobile No	9847940976
	Email	aswinihealthspecialities@gmail.com

**Item-wise Price Details**

#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Bronchoscope Cryosurgery Unit</b> <i>Model &amp; Make : Cryo 2/ Erbe Elektromedizin</i>	1936480 Incl.GST :12%	142815.4	2079295.4
		<b>1936480</b>	<b>142815.4</b>	<b>2079295.4</b>

**Annual / Comprehensive Maintenance Charges (Exl.Tax)**

Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Bronchoscope Cryosurgery Unit</b>							
Labour	35,500.00	35,500.00	35,500.00	35,500.00	35,500.00	35,500.00	35,500.00
Comprehensive	99,647.00	99,647.00	1,04,630.00	1,13,000.00	1,22,040.00	1,34,244.00	1,47,669.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Bronchoscope Cryosurgery Unit**

1. The unit should be modern cryo surgical unit with advanced features.
2. The unit should have LCD display with Plug and play socket.
3. The unit should be compatible with CO2 gas.
4. The unit should be supplied with 2 Flexible CRYO plug and play probes, Diameter - 1.9mm and Diameter 2.4mm
5. The unit should identify the instrument and set the parameters accordingly.
6. The unit should have 2 different effect settings with which we can create two different freezing effects for different Pulmonology/ Bronchoscope applications like CRYO Biopsy, CRYO De-canalizations, Tumor Debulking, Devitalizations.
7. The unit should have the following facilities:
  1. Count down, count up timers with and without alarms.
  2. Display the number of usages of CRYO probes.
  3. Digital display of available gas pressure of the CO 2 Cylinder
  4. Option to connect 2 CO 2 cylinders simultaneously
  5. Washable Foot switches
  6. Customizable programs up-to 9
  7. Updatable
8. The unit should be supplied with the following accessories:
  1. Footswitch
  2. Two gas cylinders
  3. Cart which can carry two cylinders
  4. Wire basket to carry accessories
  5. Gas hose to conduct gas from cylinder to unit
  6. CRYO flexible probe Ø 2.4mm; L 900mm
  7. CRYO flexible probe Ø 1.9mm; L 1050mm
  8. Gas bottle adapter E
  9. Switching valve with 2 Adaptor E for connecting two cylinders
  10. Exhaust gas hose; Length 10 m open