

(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647 Email :engg_ep@kmscl.kerala.gov.in CIN: U24233KL200TSGC021616, PAN : AADCK4029M, GSTIN : 32AADCK4029M1ZK

Running Contract Details			
Equipment Name	Bed Side Locker		
Running Contract Valid Till	21-03-2024		
Tender Ref No	KMSCL/EP/T422(R)/424/2021		
Tendered Quantity	5000		
Supplier Name	M/s Kerala Surgical Equipment Co		
GST No	32AACFK7295D1ZG		
Installation & Delivery Period	6 Week(s)		
Up-time / PM vist	95% & 0 Visits per year		
Warranty period	1 Years		

Supplier`s Details					
Address	Contact Details				
No.41/1737 A	Contact Person	Mr. Abbas			
MES Centre Near Town Hall Banerji Road	Phone				
Cochin -18	Mobile No	9539012460			
	Email	info@keralasurgicals.com			

Item-wise Price Details						
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total		
1	Bed Side Locker Model & Make : Medik-195	5546 Incl.GST :18%	388.22	5934.22		
		5546	388.22	5934.22		

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.

4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Bed Side Locker

- 1. overall dimensions should be 400W x 420 D x 840mm H
- 2. The Locker should have one drawer unit and one cabinet unit fitted with Cam lock.
- 3. A Gap should be provided between the drawer and Cabinet unit for storage.
- 4. The locker top should be of 0.8mm thick SS 304 Grade sheet.
- 5. The SS sheet should be finished in buff-matt finish.
- 6. The top should have raised border of 10mm height on three sides.
- 7. Drawer should be made 1 mm thick CRCA sheet and it should be fitted with ball slides for smooth drawer movement and also a recess to serve as handle.
- 8. Cabinet should made 1 mm thick CRCA sheet, fitted with hinge door and lock and should have louvers for ventilation.
- 9. Back of the locker should be provided with vertical spacer made of nylon with uniformly distributed total load bearing capacity of 20 kg.
- 10. Test certificates of these tests should be available proof loading test, cycle tests, impact test, salt spray test, castor break test.
- 11. Should be GREENGUARD / SGS & BIFMA/ANSI certified.
- 12. The manufacturer should compliant with ISO 9001, 14001, 13485 & OHSAS 18001.