

Running Contract Details	
Equipment Name	INCUBATOR - VARIOUS TYPE
Running Contract Valid Till	03-10-2025
Tender Ref No	KMSCL/EP/T400/60/2021
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details					
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Incubator lab</b> <i>Model &amp; Make : MINS-3 &amp; MINS-4/ LABLINE</i>	200	25370 <small>Incl.GST :18%</small>	1775.9	27145.9
2	<b>Incubator 450x450x600 mm</b> <i>Model &amp; Make : MINS -3 &amp; MINS-4/LABLINE</i>	200	32332 <small>Incl.GST :18%</small>	2263.24	34595.24
			<b>57702</b>	<b>4039.14</b>	<b>61741.14</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>Incubator 450x450x600 mm</b>				
Labour	3,000.00		3,000.00	3,000.00
Comprehensive	5,000.00		6,500.00	7,500.00
<b>Incubator lab</b>				
Labour	3,000.00		3,000.00	3,000.00
Comprehensive	5,000.00		6,500.00	7,500.00

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
ve			

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Incubator lab**

1. Should be operated on 230V,50Hz single-phase AC supply , and having temperature ranging from ambient to 60°C
2. Should be double-walled with stainless steel inner chamber having a minimum of two inner stainless steel shelves with holes and powder-coated outer surface.
3. The inner chamber should be fabricated with ribs for adjusting shelves to a convenient height.
4. Should have a minimum of chamber size of (L\*B\*H) of 450\*450\*450mm.
5. Should be provided with three side heating elements.
6. Should have air circulating fan (Which can be turned ON/OFF on demand) for uniform temperature on all shelves.
7. Should have a double door with an acrylic transparent door.
8. Should provide a microprocessor-based digital temperature controller with a digital display.
9. Should have synthetic rubber gasket at the door.
10. Should have a safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid
11. Traceable calibration certificate shall be provided along with each machine

#### **Equipment :Incubator 450x450x600 mm**

#### **Incubator(450\*450\*600mm)**

1. Should be operated on 230V,50Hz single-phase AC supply and having temperature ranging from ambient to 60°C
2. Should be double-walled with stainless steel inner chamber having a minimum of two inner stainless steel shelves with holes and powder-coated outer surface.
3. The inner chamber should be fabricated with ribs for adjusting shelves to a convenient height.
4. Should have a minimum of chamber size of (L\*B\*H) of 450\*450\*600mm.
5. Should be provided with three side heating elements.
6. Should have air circulating fan (Which can be turned ON/OFF on demand) for uniform temperature on all shelves.
7. Should have a double door with an acrylic transparent door.
8. Should provide with a microprocessor-based digital temperature controller with a digital display.
9. Should have synthetic rubber gasket at the door.
10. Should have a safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be

produced along with the technical bid

11. A traceable calibration certificate shall be provided along with each machine