



Running Contract Details	
Equipment Name	Digitalised Slit Lamp with Anterior Segment Photography
Running Contract Valid Till	17-11-2025
Tender Ref No	KMSCL/EP/T494(R)/1030/2023
Tendered Quantity	10
Supplier Name	M/s Appasamy Associates
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Digitalised Slit Lamp with Anterior Segment Photography <i>Model & Make : SL-CAM ICM 7/Appasamy Associates Pvt Ltd</i>	672000 Incl.GST :12%	49560	721560
		672000	49560	721560

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Digitalised Slit Lamp with Anterior Segment Photography							
Labour	18,000.00	18,000.00	18,000.00	24,000.00	24,000.00	30,000.00	30,000.00
Comprehensive	48,000.00	48,000.00	54,000.00	54,000.00	60,000.00	60,000.00	72,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Digitalised Slit Lamp with Anterior Segment Photography

1. Should have 5 step magnifications
2. Should have slit width from at least 0 to 14 mm continuously
3. Should have slit length from at least 0 to 14 mm continuously
4. Should have a slit inclination from 0 to 180 degree
5. Should have heat absorbing, red free, cobalt blue filters
6. Should have a longitudinal movement of at least 90 mm
7. Should have a lateral movement of at least 100 mm
8. Should have a vertical movement of at least 30 mm
9. Should have a fine base movement of at least 15 mm
10. Should have a chinerest vertical movement of at least 80 mm
11. Should have a fixation target
12. Should have a good illumination using LED/ halogen lamp
13. Should have an integrated camera with minimum 5 mega pixel or more
14. Should provide with company capturing software with video recording (third party software not allowed)
15. Should operate with 200 to 240 v/50 Hz power supply
16. Should be supplied with motorized table
17. Should provide computer with a good configuration (minimum 20 inch monitor, 13 processor, 1 TB hard disk, 4Gb RAM, DVD writer, UPS etc)
18. In case of Halogen Lamp 5 spare halogen bulbs to be provided (Not required for LED).
19. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid