



Running Contract Details

Equipment Name	CENTRIFUGE - VARIOUS TYPES
Running Contract Valid Till	08-01-2026
Tender Ref No	KMSCL/EP/T512/21A/2023
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details

Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details

#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Centrifuge - 8 Wells <i>Model & Make : MEDICO PLUS with SM 8- 15ml Rotor/ M/S REMI ELEKTROTECHNIK LTD</i>	500	32975.1 Incl.GST :18%	2308.26	35283.36
2	Centrifuge - 16 Wells <i>Model & Make : R 8C PLUS with S 16-15 ml Rotor/ M/S REMI ELEKTROTECHNIK LTD</i>	500	45418.2 Incl.GST :18%	3179.27	48597.47
			78393.3	5487.53	83880.83

Annual / Comprehensive Maintenance Charges (Exl.Tax)

Rate	3 rd Year	4 th Year	5 th Year
Centrifuge - 16 Wells			
Labour	1,000.00	1,000.00	1,000.00
Comprehens ive	3,000.00	3,500.00	4,000.00
Centrifuge - 8 Wells			
Labour	1,000.00	1,000.00	1,000.00

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Comprehensive	3,000.00	3,500.00	4,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Centrifuge - 8 Wells

1. Should have a maximum speed of 5000 RPM with stepless regulator
2. Should be supplied with safety lid and lock.
3. Should have digital speedometer and timer.
4. Should have imbalance detector and automatic cutoff.
5. Should have CE and ISO13485 certification.
6. Should work on 200-240Vac 50Hz power supply.
7. Should have swing out rotor of size 8*15ml

Note:

1. L1 shall be declared individually for each item
2. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
3. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail

Equipment :Centrifuge - 16 Wells

I.Centrifuge 16 wells

1. Should have a maximum speed of 5000 RPM with stepless regulator
2. Should be supplied with safety lid and lock.
3. Should have digital speedometer and timer.
4. Should have imbalance detector and automatic cutoff.
5. Should be CE Certified.
6. Should work on 200-240Vac 50Hz power supply
7. Should have swing out rotor of size 16*15ml
8. Should be brushless type & maintenance-free motor with frequency drive
9. Should have programmable time and speed.
10. Should be low noise operation type

