

Running Contract Details	
Equipment Name	CERAMIC FURNACE
Running Contract Valid Till	23-05-2025
Tender Ref No	KMSCL/EP/T489/1622A/2023
Tendered Quantity	5
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	CERAMIC FURNACE <i>Model & Make : PROGRAMAT EP 3010 G2/VIJAI DENTAL DEPOT PVT. LTD.</i>	801810 Incl.GST :18%	56126.7	857936.7
		801810	56126.7	857936.7

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
CERAMIC FURNACE							
Labour	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00
Comprehensive	20,000.00	25,000.00	30,000.00	35,000.00	40,000.00	45,000.00	50,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment : CERAMIC FURNACE

Equipment: Ceramic Furnace

1. Pressable ceramic furnace with pressing and firing program.
2. QTK2 muffle technology with SiC bottom reflector.
3. The firing chamber may be heated up to 1200°C.
4. Vacuum may be created with a vacuum pump
5. Furnace head with firing chamber
6. Wide-screen colour display
7. Operated by membrane-sealed keypad and touch screen.
8. Should be supplied with following Accessories
 - a. Vacuum pump
 - b. Firing tray kit
 - c. Automatic temperature checking set
 - d. Cooling grid
 - e. Tongs
9. Dimension of closed furnace
 - a. Depth: 465mm
 - b. Width : 390 mm (with cooling tray)
 - c. Height: 550mm
10. Power supply 200-240 V/50-60 Hz

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid