



Running Contract Details	
Equipment Name	Automated Blood Culture System
Running Contract Valid Till	29-05-2025
Tender Ref No	KMSCL/EP/T482(R)/998/2022
Tendered Quantity	6
Supplier Name	M/s Laboratory Equipment & Stores
GST No	32AABFL6606J1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
M G Road North End Near Seematti Ernakulam - 682 035.	Contact Person	Mr. Soney Mathew
	Phone	
	Mobile No	9846066199
	Email	info@labestore.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Automated Blood Culture System <i>Model & Make : Bactec FX40/ Becton Dickinson India Pvt Ltd</i>	811840 Incl.GST :18%	56828.8	868668.8
2	Rate of one anaerobic Blood Culture Bottle(Adult)	277.3 Incl.GST :18%	19.41	296.71
3	Rate of one Aerobic Blood Culture Bottle(Adult)	277.3 Incl.GST :18%	19.41	296.71
4	Rate of one Blood Culture Bottle(Pediatric)	277.3 Incl.GST :18%	19.41	296.71
5	Rate of one Aerobic Platelet Culture QC bottle	277.3 Incl.GST :18%	19.41	296.71
6	Rate of one Anaerobic Platelet Culture QC bottle	277.3 Incl.GST :18%	19.41	296.71
7	Rate of one TB Culture bottle	395.3 Incl.GST :18%	27.67	422.97
		813621.8	56953.53	870575.33

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Automated Blood Culture System							
Labour	70,800.00	70,800.00	70,800.00	70,800.00	70,800.00	70,800.00	70,800.00
Comprehensive	1,06,200.00	1,06,200.00	1,06,200.00	1,06,200.00	1,06,200.00	1,06,200.00	1,06,200.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Automated Blood Culture System

1. The system should be a fully automated, walk away system capable of culture and detection of bacteria, mycobacteria and fungi from blood and sterile body fluids.
2. Should have capacity to hold at least 40 bottles at a time. The capacity should be upgradable.
3. The system should continuously monitor the samples for growth and report it as and when it occurs.
4. The culture media provided should have sufficient mechanism to neutralise the Inhibitory effect of antibiotics and other substances in blood.
5. a) Culture media should be available for detecting bacteria and fungi, including fastidious organisms., b)The culture bottles should be unbreakable in normal conditions., C) The culture system should be suitable for processing blood and sterile body fluids.
6. Should be capable of processing both adult and the paediatric samples.
7. The system should be maintenance free without any need for regular calibrations, controls or standards run by the user.
8. The system should use leak proof and on noninvasive system to avoid contamination of equipment and environment.
9. The culture bottles should have high stability and (4-6) months shelf life.
10. The system should have all the facilities for data management and storage and quality control.
11. The system should be supplied in a complete system with all accessories, hardwarelike computer, printer, UPS etc and required software.
12. Any software or database updates should be done free of cost by the firm, during the life of equipment, as and when it is released by the manufacturer.
13. Required training, technical literature and support should be provided by the firm.
14. Any calibration, routine maintenance, and replacement of the parts like sensors, lamps etc required during the warranty and AMC and which is not covered by the same, should be declared and the respective costs quoted.
15. Should give 200 Culture bottles (Aerobic/Pediatric) as Free of Cost along with the instrument