



Running Contract Details	
Equipment Name	Pulse Oxymeter Model A
Running Contract Valid Till	30-01-2025
Tender Ref No	KMSCL/EP/T443/105A/2022
Tendered Quantity	250
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Pulse Oxymeter Model A <i>Model & Make : YONKER 810C / Xuzhou Yongkang Electronic Science Technology Co., Ltd.</i>	27988.8 Incl.GST :12%	2064.17	30052.97
2	Adult reusable Spo2 Probe	2788.8 Incl.GST :12%	205.67	2994.47
3	One forehead disposable SPO2 sensor	2788.8 Incl.GST :12%	205.67	2994.47
		33566.4	2475.52	36041.92

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
Pulse Oxymeter Model A			
Labour	500.00	500.00	500.00
Comprehensive	1,500.00	1,500.00	1,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Pulse Oxymeter Model A

1. Should have plethysmographic wave form with numeric display for SPO2 and Heart rate on LED graphic display/ LCD/TFT display screen.
2. Should have a SPO2 range of 0 to 100%.
3. Should have SPO2 accuracy of $\pm 2\%$.
4. Should provide bar graph for pulse strength.
5. Audio and visual alarm for both upper and lower SPO2, Heart rate.
6. Should quote rate separately for Reusable Adult Probe.

The rates offered will be taken for evaluation. (Rate will be fixed for 3 years)
7. Beep sound and alarm sound should have separate volume control
8. Should have a minimum of 4 hours back-up time.
9. Should be a portable, light weight and desktop model with adult, pediatric and Neonatal modes.
10. Should work with input 200 to 240Vac 50 Hz supply.
11. Should have trend data of at least 24 hrs.
12. Should provide with adult reusable finger probe with technology from standard reputed companies like Massimo, Nellcore or equivalent
13. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid
14. Rate of one forehead disposable SPO2 sensor should be quoted separately which will be taken for evaluation.