



Running Contract Details	
Equipment Name	Otoscope
Running Contract Valid Till	26-07-2025
Tender Ref No	KMSCL/EP/T494(R)/94/2023
Tendered Quantity	1500
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Otoscope <i>Model & Make : 3.6 V KEELER STANDARD OTOSCOPE (1518-P-1002)/ KEELER LTD UK</i>	27888 Incl.GST :12%	2056.74	29944.74
		27888	2056.74	29944.74

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Otoscope

1. Should be a C-Cell type or D-cell type Otoscope suitable for usage in outpatient departments of Govt. Hospitals.
2. Should be provided with a halogen/LED light source. In case of halogen light source, 2 Nos of spare bulbs are to be provided.
3. Should be able to detach the otoscope head.
4. Should provide no reflections and obstructions.
5. Should provide the following detachable accessories
 1. Specula (2.5mm to 5mm, 4 sizes)
6. Should have rechargeable battery. Charger for the Rechargeable battery shall be provided
7. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid