



Running Contract Details	
Equipment Name	Disposable flexible Video Bronchoscope
Running Contract Valid Till	28-01-2023
Tender Ref No	KMSCL/EP/T379/200B/2020
Tendered Quantity	100
Supplier Name	M/s Biomedical Engineering Company
GST No	32AAGFB1151K1ZV
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
39/878 A2 - YMJ West Lane Palarivattom Kochi 682025	Contact Person	Mr. Ajith George Mathew
	Phone	
	Mobile No	8921065594, 9447012546
	Email	mail@behealthcare.com,accounts@behealthcare.com,sales@behealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Monitor and other accesoriesfor flexible video Bronchoscope <i>Model & Make : Ambu aScope4 broncho:/Ambu</i>	451248 Incl.GST :12%	33279.54	484527.54
2	Disposable Bronchoscope of Outer diameter of 3.8 mm & inner diameter of 1.2 mm	40320 Incl.GST :12%	2973.6	43293.6
3	Disposable Bronchoscope of Outer diameter of 5 mm & inner diameter of 2 mm	40320 Incl.GST :12%	2973.6	43293.6
4	Disposable Bronchoscope of Outer diameter of 5.8 mm & inner diameter of 2.8 mm	40320 Incl.GST :12%	2973.6	43293.6
		572208	42200.34	614408.34

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Disposable flexible Video Bronchoscope							
Labour	22,576.00	24,834.00	27,317.00	30,049.00	33,054.00	36,359.00	39,995.00
Comprehensive	45,152.00	49,667.00	54,634.00	60,098.00	66,107.00	72,718.00	79,990.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Disposable flexible Video Bronchoscope

1. Equipment should be disposable flexible, light weight handy and portable bronchoscope with inbuilt light source for highly infected patients
2. Scope should be available in following various sizes.
 - a. Outer diameter of 3.8 mm & inner diameter of 1.2 mm
 - b. Outer diameter of 5 mm & inner diameter of 2.2 mm
 - c. Outer diameter of 5.8 mm & inner diameter of 2.8 mm
3. Working length 600 mm
4. Should have suction port, oxygen port and drug installation port
5. Tip should be easily manoeuvrable, tip deflection range 180/180 degree
6. Scope should come in pre sterile packing
7. Should provide portable TFT LCD monitor of size at least 8" with touch screen and clamp for fixing with IV stand.
8. Monitor should have minimum 8 GB storage capacity, minimum 2 hour backup, recording facility, USB port and provision for video output
9. Monitor should run on both electricity and battery
10. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid. (for monitor).

