



Running Contract Details

| | |
|--------------------------------|-------------------------|
| Equipment Name | Gynec OT Table |
| Running Contract Valid Till | 11-01-2023 |
| Tender Ref No | KMSCL/EP/T351/293L/2019 |
| Tendered Quantity | 100 |
| Supplier Name | M/s Trichur Surgicals |
| GST No | 32AAFFT9094G1ZX |
| Installation & Delivery Period | 8 Week(s) |
| Up-time / PM vist | 95% & 4 Visits per year |
| Warranty period | 3 Years |

Supplier`s Details

| Address | Contact Details | |
|----------------------------------|-----------------|-----------------------------|
| Veliyannur Road Thrissur - 21 | Contact Person | CD David |
| | Phone | |
| | Mobile No | 9447020813 |
| | Email | trisco@trichursurgicals.com |

Item-wise Price Details

| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
|---|--|---|------------------------------------|---------------|
| 1 | Gynec OT Table <i>Model & Make : TS115/Trichur Surgicals</i> | 283200 Incl.GST :18% | 19824 | 303024 |
| | | 283200 | 19824 | 303024 |

Annual / Comprehensive Maintenance Charges (Exl.Tax)

| Rate | 4 th Year | 5 th Year | 6 th Year | 7 th Year | 8 th Year | 9 th Year | 10 th Year |
|-----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Gynec OT Table | | | | | | | |
| Labour | 5,750.00 | 6,900.00 | 8,050.00 | 9,200.00 | 10,350.00 | 11,500.00 | 12,650.00 |
| Comprehensive | 10,810.00 | 11,960.00 | 13,110.00 | 14,260.00 | 15,410.00 | 16,560.00 | 17,710.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Gynec OT Table

1. The table should be electrically operated.
2. It should have overall approximate size 1830 (l) x 560 (w)mm without accessories
3. The table top should have a minimum height of 680mm, vertical stroke of minimum of 300mm
4. Should have at least 25° trendelenburg and reverse trendelenburg
5. Should have at least 15° lateral tilt movements
6. Should have at back up movement with gas spring mechanism of 65° up/ 25° down.
7. Should have head rest, hand rest and leg rest detachable with fixing clamps.
8. Should have SS/ABS side rails with safety guard for clamps on both sides of the table, column covers the electro mechanical actuator system
9. Table top should be fitted with X-ray permeable top
10. The table should have heavy metal base covered with molded ABS/SS.
11. The table should be fitted with non rusting 50mm dia casters, foot lever operated table lock
12. All MS parts used in the table should be pretreated and epoxy powder coated
13. Should be supplied with one pair of leg crutches with side rail locks, shoulder rest (one pair) with side rail locks, side support with lock.
14. Cushions' mountable and 3 pieces- head, body and foot with minimal gap