



Running Contract Details	
Equipment Name	Leg Stirrups with Lift Assist for Lithotomy and Abduction
Running Contract Valid Till	29-12-2022
Tender Ref No	KMSCL/EP/T362/1522/2020
Tendered Quantity	10
Supplier Name	M/s Champak Enterprises Pvt. Ltd
GST No	29AABCC3219R1Z9
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Champak No : 29 - 5th Main Tata Silk Farm - Basavangudi Bangalore 560004 Karnataka	Contact Person	Mr. V. Pattabhiraman
	Phone	
	Mobile No	9844073940
	Email	champakmed@vsnl.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Leg Stirrups with Lift Assist for Lithotomy and Abduction <i>Model & Make : Magna Fins/Magnatek</i>	660800 Incl.GST :18%	46256	707056
		660800	46256	707056

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Leg Stirrups with Lift Assist for Lithotomy and Abduction							
Labour	7,200.00	8,000.00	8,889.00	9,878.00	10,973.00	12,193.00	13,549.00
Comprehensive	48,000.00	53,334.00	59,260.00	65,844.00	73,160.00	81,289.00	90,321.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Leg Stirrups with Lift Assist for Lithotomy and Abduction

1. System should be specialized leg stirrups with lift assist technology
2. Product should be of international quality with CE and ISO certified
3. Yellow fin stirrups with lift assist for lithotomy and abduction with weight carrying capacity of 155 Kg or above
4. The abduction range should be between +25 degree to -09 degree
5. Lithotomy positioning range should be possible between +84 degree to -33 degree
6. Extending the boots to the lateral side (left and right) should be possible by squeezing the grip handles
7. Leg holder should be moveable in all angled directions by squeezing the grip handle
8. Indicators/ scales marking should be available for precise positioning
9. Raise and abduction by single hand for ease of operation
10. The angle of the pad lining of the boots should be in such a way that it should cover the peroneal nerve, besides covering the head of the fibula
11. Boot should be self adjusting type to minimize pressure on the calf
12. Should be supported with required reusable pads and clamps for attachment
13. Should supply specialized trolley to hang the stirrups when not in use