



**Running Contract Details**

Equipment Name	RF Cautery
Running Contract Valid Till	10-01-2023
Tender Ref No	KMSCL/EP/T356(R)/763/2019
Tendered Quantity	15
Supplier Name	M/s C Cube Advanced Technologies
GST No	29ANCPR4091F1ZS
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

**Supplier`s Details**

Address	Contact Details	
406 3rd floor Malnad Residency 4th Main Hanuma Hills layout Banashankari 3rd stage Bangalore 560061	Contact Person	Vinayaka KS
	Phone	8043021028
	Mobile No	9611111273
	Email	ccubeadvtech@gmail.com

**Item-wise Price Details**

#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>RF Cautery</b> <i>Model &amp; Make : RF Surg/C CUBE</i>	392000 Incl.GST :12%	28910	420910
		<b>392000</b>	<b>28910</b>	<b>420910</b>

**Annual / Comprehensive Maintenance Charges (Exl.Tax)**

Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>RF Cautery</b>							
Labour	15,000.00	15,000.00	17,500.00	17,500.00	17,500.00	22,000.00	22,000.00
Comprehensive	20,000.00	20,000.00	20,000.00	20,000.00	30,000.00	30,000.00	35,000.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :RF Cautery**

1. Should required dual frequency (monopolar+ bipolar)
2. Output frequency should be minimum of 3.8 MHz
3. Should have reusable hand pieces and antenna plate
4. The following modes should be available - cutting, coagulation, fulgarisation
5. System should have minimum lateral tissue damage
6. Should have multiple electrodes of different shapes