



Running Contract Details	
Equipment Name	Syringe Infusion Pump (Dual)
Running Contract Valid Till	14-02-2022
Tender Ref No	KMSCL/EP/T346(R)/1041/2019
Tendered Quantity	100
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Syringe Infusion Pump (Dual) <i>Model & Make : Benefusion SP3D/ Shenzhen Shenke</i>	66024 Incl.GST :12%	4869.27	70893.27
		66024	4869.27	70893.27

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Syringe Infusion Pump (Dual)				
Labour	2,500.00	2,500.00		2,500.00
Comprehensive	5,895.00	6,190.00		6,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Syringe Infusion Pump (Dual)

1. Two syringe drive system can deliver drugs independently and simultaneously
2. Pressure controlled mechanism
3. 3" LCD display or more
4. Should have minimum flow rate range from 0.1 -1200ml/hr for 50ml syringe, 0.1 -100ml/hr for 20ml syringe and 0.1-60ml for 10ml syringe
5. Bolus rate : maximum 1000ml/hr
6. KVO facility & 50 drug library
7. Should have automatic syringe size and model detection.
8. 3 level alarm pressure settings
9. Volume and time controlled infusion
10. Mains and battery operated
11. Easy to use interface
12. Should have automatic and manual bolus
13. Should have atleast 3 levels of programmable occlusion pressure
14. Should have a rechargeable Lithium ion battery with backup time of minimum 3 hours.
15. Should have automatic bolus reduction system to avoid accidental bolus delivery after occlusion incident.
16. Syringe ranges from (5-60) ml
17. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the

technical bid