



Running Contract Details	
Equipment Name	HOSPITAL FURNITURE
Running Contract Valid Till	10-01-2021
Tender Ref No	KMSCL/EP/T293/551/2018
Tendered Quantity	500
Supplier Name	M/s SP Pharma Distributors
GST No	32ACCFS4349E1ZD
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
32/2964 Near St. Mathews Church Vennala P.O. Palarivattom Ernakulam - 682028	Contact Person	Mr.Abraham George
	Phone	0484 - 4012447
	Mobile No	9895522447
	Email	sp.surgicals@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Emergency Trolley (Trauma Emergency Cot) <i>Model & Make : Mi-4076 A / Medimek Industries</i>	47200 Incl.GST :18%	3304	50504
2	Patient transfer roller <i>Model & Make : Mi-1044 A / Medimek Industries</i>	5550.72 Incl.GST :18%	388.55	5939.27
3	Pigeon hole locker cupboard with 12 locker <i>Model & Make : Mi-3021 A / Medimek Industries</i>	26432 Incl.GST :18%	1850.24	28282.24
		79182.72	5542.79	84725.51

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Emergency Trolley (Trauma Emergency Cot)

1. Should have two sectional top X ray translucent high pressure laminate with facility to insert X-ray cassette from both sides and ends of trolley.
2. Should be able to X ray the patient from positions the entire length and width of the trolley
3. Should have ratchet adjusted back section
4. Should have trendelenberg and reverse trendelenberg positions.
5. Should have hydraulic height adjustment.
6. Should have removable stretcher for easy cleaning and disinfections of the X ray platform.
7. Should have place for fixing oxygen cylinder.
8. Should have detachable SS telescopic IV rods and collapsible SS side rails.
9. Body should be pre-treated and power coated.
10. Should be supplied with standard accessories such as
 - a. SS Collapsible Side rails - 1 no.
 - b. Height adjustable IV rod - 1 no.
11. Dimension:-
 - a. Max.Length : 1900±20 mm
 - b. Max.Width : 700±10 mm
 - c. Height : Min 65 ±5 cm Max 90 ±5 cm
 - d. Trendelenberg : 10 ±2 deg.
 - e. Reverse trendelenberg : 5 ±1 deg
 - f. X ray viewing area Entire length
12. The manufacturer should compliant with ISO 9001, 14001, 13485 & OHSAS 18001.
13. Should be provided with mattress of 32 density foam 2” thick with two way stretchable rexine.

Equipment :Patient transfer roller

1. Board Dimensions shall be 15" W x 30" L x 1.5" H,

2. Should have at least five rollers of non corrosive, non rusting material
3. Easy slide of patient along the covered ball bearings to the required position
4. Should be provided with one cover of 2mm thick lectrolite material.
5. Patient handling capacity at least 150 kg

Equipment :Pigeon hole locker cupboard with 12 locker

1. Steel locker almirah – 78”x36”x19”
2. 12 lockers (3 lockers in row x4) with separate lock.
3. The thickness of the sheets (shelf, back, sides and door) shall be at least 20 gauge CR sheet.
4. Pre-treated and powder coated.
5. It should have welded construction.
6. The components shall be made of CRCA sheet.
7. The finished product should be Epoxy powder coated to the thickness of 50 microns (+/- 10).
8. The manufacturer should compliant with ISO 9001, 14001, 13485 & OHSAS 18001