



Running Contract Details	
Equipment Name	Video Colposcope HD Type II
Running Contract Valid Till	15-08-2021
Tender Ref No	KMSCL/EP/T333/909B/2019
Tendered Quantity	16
Supplier Name	M/s Inskon Medical Solutions Private Limited
GST No	32AACCI7233A1Z8
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
98C Manthrakkal Junction Thaikkattukara Post Aluva Kochi- 683106	Contact Person	ajesh
	Phone	
	Mobile No	9744160009
	Email	sales@inscon.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Video Colposcope HD Type II <i>Model & Make : Leisegang 3MVC/3MVCS LED / Cooper Surgicals, USA</i>	907200 Incl.GST :12%	66906	974106
2	Computer / laptop with accessories	165200 Incl.GST :18%	11564	176764
		1072400	78470	1150870

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Video Colposcope HD Type II							
Labour	30,000.00	32,500.00	35,000.00	37,500.00	40,000.00	42,500.00	45,000.00
Comprehensive	70,000.00	75,000.00	80,000.00	85,000.00	90,000.00	95,000.00	1,00,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment : Video Colposcope HD Type II

1	The output of video colposcope should be full HD (1920*1080)
2	Should have maintenance free white LED light source-(6000 lux) with a recommended lamp life of 20,000 working hours
3	Should have fast auto/ Manual focusing, Fine focus, Zooming, Image Freeze, Image capture control, Green Filter can be either electronic or mechanical and timer all in a handheld Unit
4	Should have internal image acquisition control
5	Should have anti- vibration Image and should ensure clear image under high magnification
6	Image output: S- Video, Pal, Ntsc, HDMI, Y/Pb/Pr output
7	Working Distance of 200-300mm
8	1-50X continuous ZOOM
9	Field of view: 5-50mm
10	Depth of view: 5mm-120mm
11	Should show display of true tissue colour and tone rendition without loss of any magnification
12	Imaging System with 12-15+ Megapixel
13	Resolution of >850 lines
14	Should have latest Digital Matrix Processor Technology 1920x1080 P
15	LED light source fitted on a vertical height adjustable floor stand, manufactured by the OEM company (height adjustment >=150 mm)
16	Equipment to be supplied with complete data management Systems (with original software)
17	Equipment to be supplied with original workstation (Preferably laptop)
18	Should provide Insulated end cervical Speculum and lateral Vaginal Wall retractor
19	Should provide two numbers of Imported explosion proof footswitch for use with workstation to save the pictures.
	Multi-function Image processing software- specifications
20	Management of examination viewing, Image capture, Image Freeze, Video Recording, Video editing and recall of archived images and videos.
21	Integration to LAN and HIS when required
22	Should have the facility to point and mark the biopsy site
23	Adequate storage for colposcopy pictures
24	Reference pictures and videos inbuilt in the software
25	Printing of multi-format diagnosis report like four photo report
26	Facility to transfer data to DVD, USB and email
27	Colposcopy SWEDE score Evaluation
	Computer / Laptop Specification for HD Video Colposcope
28	Hard Disk 1 TB
29	RAM 4 GB
30	Processor CORE i5
31	HD Monitor
32	HDMI Output Card (Specially configured for HD Colposcope)

33	Key Board 85 Mouse
34	UPS of suitable capacity with 30 minutes backup (not required if laptop).
35	HP ink-tank Color Printer for reporting
	Safety Standards and Warranty
36	Should have safety certificate from a competent authority CE issued by a notified body registered in European commission CE/ FDA (US)/ISO
37	Comprehensive training for staff and support services till familiarity with the system.