



Running Contract Details	
Equipment Name	Video Laryngoscope Model B
Running Contract Valid Till	01-11-2021
Tender Ref No	KMSCL/EP/T333/199B/2019(R)
Tendered Quantity	200
Supplier Name	M/s India Medtronic Pvt Ltd
GST No	27AAACI4227Q1Z8
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
1241 Solitaire Corporate Park Building No.12 4th Floor Andheri-Ghatkopar Link Road Andheri - East Mumbai - 400093	Contact Person	Pradeep Kumar S G
	Phone	02233074700
	Mobile No	9884966277
	Email	s.g.pradeep.kumar@medtronic.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Video Laryngoscope Model B <i>Model & Make : McGrath MAC / Medtronic</i>	168000 Incl.GST :12%	12390	180390
2	size 1 blade	616 Incl.GST :12%	45.43	661.43
3	size 2 blade	616 Incl.GST :12%	45.43	661.43
4	size 3 blade	616 Incl.GST :12%	45.43	661.43
5	size 4 blade	616 Incl.GST :12%	45.43	661.43
6	Blade for difficult Intubation	1120 Incl.GST :12%	82.6	1202.6
7	battery McGrath Mac 3.6V	2242 Incl.GST :18%	156.94	2398.94

Item-wise Price Details			
		173826	12811.26
			186637.26

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Video Laryngoscope Model B							
Labour	1,660.00	1,820.00	2,010.00	2,200.00	2,400.00	2,640.00	2,900.00
Comprehensive	5,000.00	5,490.00	6,040.00	6,640.00	7,300.00	8,030.00	8,830.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Video Laryngoscope Model B

1. Should be a portable video laryngoscope for intubations with minimal manipulation of head & neck dedicated features for teaching, training & learning in the specialty
2. Minimum 1 megapixel camera should be available
3. Should have a free fog optical polymer material/ poly carbonate material blades
4. Should have a suitable view angle to visualize glottis without much head & neck manipulation, ergonomically
5. The system should have portable colour video display LCD of at least 2.3 ” or above size for the real time clear view
6. Weight of handle should be light and not be more than 250 g
7. Should have passed the drop test for one meters
8. Light sources should be high-intensity LED
9. Should have facility to run independently on a battery and back up should have minimum four hours.

The rate for the battery shall be offered in the BOQ and the same will be fixed for 5 years from the date of price bid opening. The rate will be taken for evaluation.

10. The system should be supplied with a set of different sizes of disposable blade size 1,2,3,4 and one additional blade for difficult intubation
11. Should be immersible for complete disinfection (without battery)
12. Should supply the following blades free of cost along with the machine - 25 nos. of size 1, size 2, size 3, size 4 and 10 numbers of 'additional blade for difficult intubation'
13. The rate for the 5 types of blades shall be mentioned in the BOQ (taken for evaluation) and the rate will be freeze for 5 years from the date of price bid opening.
14. Device should have durable medical grade thermoplastics
15. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US). Copy of the certificate / test report shall be produced along with the technical bid