



Running Contract Details	
Equipment Name	Binocular Microscope - Type III
Running Contract Valid Till	24-11-2024
Tender Ref No	KMSCL/EP/T448/13C/2022
Tendered Quantity	800
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Binocular Microscope - Type III</b> <i>Model &amp; Make : VISION 2000 / N. K. JAIN INSTRUMENTS PVT. LTD.</i>	22939.2 Incl.GST :18%	1605.74	24544.94
		<b>22939.2</b>	<b>1605.74</b>	<b>24544.94</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Binocular Microscope - Type III**

1. Inter changeable, inclined Binocular body, 360 degree rotatable head.
2. Eyepieces - Highest quality 10 X wide field anti fungus field eyepiece.
3. Objectives- Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with semiplan achromatic correction for FOV of 18 mm.
4. Sub stage - Abbe condenser N.A 1.25 focusable, continuously variable iris diaphragm
5. Illuminator - Built-in LED/ Halogen light source with white light.
6. Finish - A durable textured acid resistant finish.
7. Should provide dust cover and immersion oil.
8. Should have Electrical safety certification
9. Should work with input 200 to 240Vac 50 Hz supply