



Running Contract Details	
Equipment Name	Point of care ABG machine
Running Contract Valid Till	21-07-2024
Tender Ref No	KMSCL/EP/T399/17A/2021
Tendered Quantity	100
Supplier Name	M/s Cyrix Healthcare Pvt Ltd
GST No	32AAFCC2499H2ZM
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Deepam - VII/287 Chitrapuzha Road Thiruvankulam Kerala - 682 305	Contact Person	Mr. Ajithkumar S
	Phone	0484-6544033
	Mobile No	
	Email	finance.cyrix@gmail.com, acc.kl@cyrix.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Point of care ABG machine</b> <i>Model &amp; Make : epoc Siemens Healthcare Pvt</i>	302080 Incl.GST :18%	21145.6	323225.6
2	Catridge 1(Test of Glucose,Na,K,Chloride,Ionised Calcium,HCT,PO2 and PCO2,Lactate and Creatinine)	471.52 Incl.GST :12%	34.77	506.29
		<b>302551.52</b>	<b>21180.37</b>	<b>323731.89</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Point of care ABG machine</b>							
Labour	6,000.00	6,000.00	6,000.00	6,000.00	6,000.00	6,000.00	6,000.00
Comprehensive	21,000.00	21,000.00	21,000.00	21,000.00	21,000.00	21,000.00	21,000.00

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Point of care ABG machine**

#### **Point of Care ABG Machine**

#### **Portable single handheld system for blood gases, lactate, hematology, chemistries, electrolytes, coagulation and cardiac markers**

1. A single portable handheld system weight of the device is less than 1 Kg capable of measuring blood gas, lactate, and other parameters as mentioned below.
2. It should give the results of GLUCOSE, Na, K, CHLORIDE, IONISED CALCIUM, HCT, PO<sub>2</sub> and PCO<sub>2</sub>, Lactate, and Creatinine. These tests should be done using not more than 5 cartridges. Each test should be provided with calibrators. The price of the cartridge will be taken for evaluation
3. It should have at least 250 test result storage facilities.
4. It should have a facility to take printouts.
5. Should have a safety certificate from a competent authority CE issued by a notified body registered in the European commission /FDA(US)/STQC CB certificate /STQC S certificate or a valid detailed electrical and functional safety test report from ERTL. Copy of the certificate /test report shall be produced along with the technical bid.
6. The system should be operated either through the rechargeable battery or 230Vac with battery backup. In case of rechargeable battery, an additional one set battery and charger to be provided
7. The cartridge should have at least 3 months of shelf life from the date of supply of hospitals. The bidder shall also replace the cartridges nearing expiry at free of cost on intimation from the hospital.