

Running Contract Details	
Equipment Name	Microwave Diathermy
Running Contract Valid Till	05-01-2025
Tender Ref No	KMSCL/EP/T463/273/2022
Tendered Quantity	25
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Microwave Diathermy <i>Model & Make : IEMCO250CP/ International Electro Medical Co</i>	551880 Incl.GST :12%	40701.15	592581.15
		551880	40701.15	592581.15

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Microwave Diathermy				
Labour	14,785.00	14,785.00		14,785.00
Comprehensive	34,500.00	34,500.00		34,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Microwave Diathermy

1. Microwave diathermy is a therapeutical method that applies electromagnetic energy in microwave frequency bands and is intended to generate deep heat within body tissues.
2. The application of microwave diathermy increases blood flow, speeds up metabolism and the rate of ion diffusion across cellular membranes.
3. It should be set up on a single click, no assembly required.
4. There should be unlimited 360 deg rotation in 2 independent joints.
5. There should be maximum position variability.
6. There should be a universal spot radiator, large field radiator, long radiator and special contact applicator .
7. There should be 5.7'' touch screen display LED/LCD user interface.
8. The power should be 250W for optimal application possibilities
9. There should be continuous and pulsed modes
10. The mains supply should be 220-240V ,50-60 Hz
11. RF table for the radiators, mains power cable and user manual.
12. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL