

(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647 Email :ep.kmscl@kerala.gov.in CIN: U24233KL200TSGC021616, PAN : AADCK4029M, GSTIN : 32AADCK4029M1ZK

Running Contract Details			
Equipment Name	Holter Monitor Model C		
Running Contract Valid Till	07-08-2025		
Tender Ref No	KMSCL/EP/T489(R)/160C/2023		
Tendered Quantity	5		
Supplier Name	M/s Schiller Healthcare India Pvt. Ltd		
GST No	34AADCS5091B1ZN		
Installation & Delivery Period	8 Week(s)		
Up-time / PM vist	95% & 4 Visits per year		
Warranty period	3 Years		

Supplier`s Details					
Address	Contact Details				
Schiller Healthcare India Pvt Ltd 37/1013 - 2nd Floor -	Contact Person	Mr.Stijo Mathew, Mr. P.S Vidhyashankar			
Unidec House Fathima Church Road Elamkulam - 682 020	Phone	04842203264			
	Mobile No	09846706787, 09383620520			
	Email	alok@schillerindia.com, Mail@schillerindia.com, shinoj@schillerindia.com, kishorekumar@schillerindia.com, stijo@schillerindia.com, vidyashankar@schillerindia.com, mail@schillerindia.com, kishorekumar@schillerindia.com			

	Item-wise Price Details							
#	Item Details					harges MSCL)	Grand Total	
1	1 Holter software Model & Make : Medilog AR/M/s. Schiller AG			221200 Incl.GST :12%		16313.5	237513.5	
2	Recorders & accessories			207090 Incl.GST :18%		14496.3	221586.3	
3	3 Computer			12956 Incl.GST :18		9069.48	138633.48	
				55785	54 3	39879.28	597733.28	
Annual / Comprehensive Maintenance Charges (Exl.Tax)								
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year	

	Annual / Comprehensive Maintenance Charges (Exl.Tax)							
	Holter software							
Labour	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	
Comprehensi ve	25,200.00	25,200.00	25,200.00	25,200.00	25,200.00	25,200.00	25,200.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.

4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Holter software

Equipment: Holter Monitor Model C

- 1. PC Supported System with following specifications
 - 1. Intel I5 processor
 - 2. Windows licensed inbuilt OS
 - 3.8 GB RAM
 - 4.1 TB HDD
 - 5. Monitor 21 inch or more
 - 6. Computer table
- 2. Saved data should be read using integrated reading device
- 3. Three channel ECG recording facility
- 4. Supporting hardware with suitable software and laser printer
- 5. Adult patients
- 6. 5 or 7 lead cable
- 7. Automatic cable type detection
- 8. Light weight
- 9. 24hrs, 48hrs, 7 days
- 10. True signal quality check with amplitude indication
- 11. QT/ST Analysis
- 12. LCD/LED Screen
- 13. Shock and splash-proof design
- 14. Suppress artifacts
- 15. Available scroll function for the over view (Super page) of the ECG
- 16. Should be able to reedit individual form classes
- 17. Should be able to consolidate individual form classes
- 18. System should be capable of analyzing various arrhythmias like ventricular ectopics, supraventricular ectopic, ventricular tachycardia, ventricular fibrillation, supraventricular tachycardia, atrial fibrillation, sinus pause
- 19. Operator should be able to edit and reclassify beats and aarhythmias

- 20. ST segment analysis should be available for all three channels
- 21. Any ECG print out should be possible
- 22. Report format should include covering page, arrhythmia analysis report, ST segment analysis, automatically and manually selected ECG strips
- 23. Should provide rechargeable battery and charger for recorders. Each recorder should be provided with 4 batteries and one charger

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid