



(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647

Email:ep.kmscl@kerala.gov.in

CIN: U24233KL200TSGC021616, PAN: AADCK4029M, GSTIN: 32AADCK4029M1ZK

Running Contract Details			
Equipment Name	High Flow Nasal cannula Therapy device		
Running Contract Valid Till	06-10-2025		
Tender Ref No	KMSCL/EP/T411/699/2021		
Tendered Quantity	150		
Supplier Name	M/s Biomedical Engineering Company		
GST No	32AAGFB1151K1ZV		
Installation & Delivery Period	6 Week(s)		
Up-time / PM vist	95% & 4 Visits per year		
Warranty period	3 Years		

Supplier`s Details					
Address	Contact Details				
39/878 A2 - YMJ West Lane	Contact Person	Mr. Ajith George Mathew			
Palarivattom Kochi 682025	Phone				
	Mobile No	8921065594, 9447012546			
	Email	mail@bechealthcare.com,accounts@bechealthcare.com,sales@bechealthcare.com			

	Item-wise Price Details					
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total		
1	High Flow Nasal cannula Therapy device  Model & Make : PTIOIUK/Fisher & Paykel Healthcare	274400 Incl.GST :12%	20237	294637		
2	Heated wire patient breathing tube	5096 Incl.GST :12%	375.83	5471.83		
3	Nasal cannula for infant	2514.4 Incl.GST :12%	185.44	2699.84		
4	Nasal cannula for peadiatric	2514.4 Incl.GST :12%	185.44	2699.84		
5	Nasal cannula for Adult	1323.84 Incl.GST :12%	97.63	1421.47		
	•	285848.64	21081.34	306929.98		

# **Other terms & conditions**

- 1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
- 2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
- 3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
- 4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
- 5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

# **Equipment : High Flow Nasal cannula Therapy device**

### **High Flow Nasal Cannula Therapy Device**

- 1. Suitable for treatment of Hypoxemic patients with respiratory distress
- 2. It should be a complaint for use on patients in ICU, wards, emergency department and home oxygen therapy
- 3. It should be a single system for treating infants, paediatric and adult patients
- 4. Inbuilt flow generator capable of delivering a wide range of flows: 2-25 litres in paediatric mode and 10-60 litres in adult mode
- 5. Inbuilt Air/O2 blending and Fio2 monitoring, facility to deliver a wide range of oxygen concentration (Fio2) from 21 to 100%
- 6. It should have an inbuilt Air source without the need for an external compressor
- 7. Integrated heated humidifier
- 8. Colour display to monitor humidity setting, flow, Fio2 and faults
- 9. Visual and audible alarm indication for:
  - a. Tubes disconnect leaks, tube blockages and water out and hardware fault with error codes. Audible power failure alarm
- 10. Disinfection mode with heated disinfection tube for sterilization of the device after patient use
- 11.Rate of one heated wire patient breathing tube and the rate of one nasal cannula of infant, peadiatric & adult should be offered separately.
- 12.Paediatric nasal cannula should be made of kink proof material and has adhesive wiggle pads to stick on skin to facilitate kangaroo care
  - 13. It should be compatible for use on tracheostomy patients
- 14. Should have a safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid Complaint with international safety standards and regulations.
  - 15. Warranty need not be provided for circuits and prongs.
  - 16. Should supply one O2 flowmeter with a capacity of 30 60 LPM
  - 17. One High Flow Flowmeter of 50 litre per minute capacity along with each unit shall be provided