



Running Contract Details	
Equipment Name	ECG Simulator
Running Contract Valid Till	22-03-2023
Tender Ref No	KMSCL/EP/T375(R)/892/2020
Tendered Quantity	20
Supplier Name	M/s BNS HEALTH AIDS
GST No	07BHYP9311L1ZD
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
3363 2ND FLOOR MAHINDRA PARK OPP. ANDHRA BANK RANI BAGH DELHI - 110034	Contact Person	RAJESH KUMAR SHARMA
	Phone	
	Mobile No	9999033454
	Email	bnshealthdelhi@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	ECG Simulator <i>Model & Make : Heartism/Laerdal Medical</i>	127079.42 Incl.GST :18%	8895.56	135974.98
		127079.42	8895.56	135974.98

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
ECG Simulator				
Labour	8,056.00	8,056.00		8,056.00
Comprehensive	4,028.00	4,028.00		4,028.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :ECG Simulator

1. Simulator is Battery Powered designed to provide basic and modified adult rhythm as well as pediatric rhythms.
2. Have a wide range of ECG rhythms as per ACLS / PALS
3. Should demonstrate Torsade de Pointes
4. Should have Infant rhythms as per PALS.
5. Should have paroxysmal; ignore the shock and variable pulse strength features.
6. It has to be connected to both normal ECG monitors and compatible manikins.
7. It is supplied with a soft case.
8. It should be compactable with standard monitors