



Running Contract Details	
Equipment Name	Ophthalmic chair unit (Refraction Unit)
Running Contract Valid Till	03-06-2026
Tender Ref No	KMSCL/EP/T519/726/2023
Tendered Quantity	50
Supplier Name	M/s LUMEN MEDIWORKS PVT LTD
GST No	32AAECL4197J1ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
63/1066A1 AISWARYAMRUTHAM CHEMMATH ROAD GANDHI NAGAR-682017	Contact Person	ARAVIND SUKUMAR
	Phone	0484-3551448
	Mobile No	6238826591
	Email	info@lumenmediworks.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Ophthalmic chair unit (Refraction Unit) <i>Model & Make : MR2012/TWOM OPHTHOTRONICS</i>	106000.16 Incl.GST :12%	7817.51	113817.67			
		106000.16	7817.51	113817.67			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Ophthalmic chair unit (Refraction Unit)							
Labour	2,900.00	2,900.00	2,900.00	2,900.00	2,900.00	2,900.00	2,900.00
Comprehensive	9,850.00	9,850.00	11,600.00	11,600.00	13,050.00	14,500.00	14,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Ophthalmic chair unit (Refraction Unit)

â€œ

Equipment: Ophthalmic Chair Unit

1. Should have microcontroller or microprocessor based control and operating system
 2. Should have controls from multiple locations
 3. All movement of chair should be adjusted with control switches
 4. Should have chair zero position
 5. Should have telescopic drawer for trial lens set
 6. Should have minimum two instrument table
 7. Should have side top with telescopic drawer
 8. Should have ophthalmoscope stand and indirect ophthalmoscope stand
 9. Should have foot switch for chair adjustment
 10. Should have head rest on chair
 11. Chair rotating should be 0- 180 degree
 12. Chair reclining should be 95 – 170 degree
 13. Seat height should be UP movement of 710mm and above and down movement of 520mm and above
 14. Weight carrying capacity up to 150 +/- 10 % Kg
 15. Chair unit should be available with left and right hand side
 16. Chair unit available with different colours
 17. Should supply the following accessories along with refraction unit
 - A. Pneumatic chair for doctor.
 - B. Powder coated metal Rim Trial lens set
 - C. LED vision chart (45.7 cm or higher))
1. Should be operated with 230V/ 50 Hz supply
 2. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid
 3. Should have adjustable leg support