



Running Contract Details	
Equipment Name	Intra Oral digital radiography system
Running Contract Valid Till	05-12-2024
Tender Ref No	KMSCL/EP/T446(R)/640B/2022
Tendered Quantity	10
Supplier Name	M/s Dentlines
GST No	32ADHPM5655K1ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Adam Tower M.C.Road Kottayam Kerala	Contact Person	MANI MATHEW
	Phone	2582506
	Mobile No	9447177082
	Email	dentlines@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Intra Oral digital radiography system <i>Model & Make : VISTASCAN MINI PLUS / DURR DENTAL</i>	644000 Incl.GST :12%	47495	691495
		644000	47495	691495

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Intra Oral digital radiography system							
Labour	33,500.00	36,500.00	40,000.00	44,500.00	49,000.00	54,000.00	60,000.00
Comprehensive	75,000.00	1,00,000.00	1,25,000.00	1,50,000.00	1,75,000.00	2,00,000.00	2,25,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Intra Oral digital radiography system

1. Should be able to use along with the existing X-ray system in the department.
2. Should have the following component
 - i. Imaging plates (size 0-4).
 - ii. Scanner for the imaging plate.
 - iii. Following accessories should be supplied along with the equipment free of cost: Imaging plates, Size 0 - 2nos, Size 1 – 1 Nos, size 2 – 6 nos, Size 3 – 1 Nos, Size 4- 2 nos
3. Specification of Imaging Plate
 - i. Should be flexible and thin as film for simple positioning.
 - ii. Cable- free system
 - iii. Sizes: 5 sizes should be available such as: size 0: 2x3cm, size 1: 2 x 4 cm, Size 2: 3.1 x 4.1 cm, Size 3: 2.7 x 5.4 cm, size 4: 5.7x7.5cm
 - iv. Broader exposure time frame & wide grey scale.
4. Specification for scanner:
 - i. Should have input/output buffer; with its cassette transport system, Table top type.
 - ii. Scanning time should be less than 1 minute