



Running Contract Details	
Equipment Name	Binocular Microscope LED with Camera
Running Contract Valid Till	28-04-2026
Tender Ref No	KMSCL/EP/T517(R)/13A3/2023
Tendered Quantity	10
Supplier Name	M/s Radical Scientific Equipments Pvt Ltd
GST No	06AACCR8985N1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
9th Milestone Ambala-Jagadhri Road Khudda Kalan Ambala Cantt-133104 (Haryana)	Contact Person	Mr. Vikram Singh Chauhan
	Phone	
	Mobile No	89030997827
	Email	tenders@radicalscientific.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Binocular Microscope LED with Camera <i>Model & Make : RXLr-4 Code-Ex/Radical Scientific Equipment Pvt. Ltd.</i>	146320 Incl.GST :18%	10242.4	156562.4
		146320	10242.4	156562.4

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Binocular Microscope LED with Camera				
Labour	2,480.00	2,480.00		2,480.00
Comprehensive	4,960.00	4,960.00		4,960.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Binocular Microscope LED with Camera

1. Should have an inclined trinocular tube with 30 degswiveling eyepiece tube or a binocular head with in-built camera.
2. Should have infinity corrected optics with anti-fungal treatment.
3. Should have built in long lasting led illumination and battery backup.
4. Should have wide field plan corrected 10X eyepiece of minimum 23mm FOV with diopter correction.
5. Should have a 5 position nose piece with 2.5x, 4x, 10x, 40x and 100x, all full plan corrected objectives.
6. All objectives should be parfocal and perfectly centered.
7. Slide stage should have ball bearing specimen holder and should have right handed coaxial X and Y movement with travelling range of min. 75 x 25 mm.
8. The condenser should have centering controls.
9. Should have a variable field diaphragm.
10. Should have an in-built camera or one attached to the trinocular tube.
11. The camera should have CMOS sensor, size 1/2.8" or better, HDMI / USB connectivity, capable of 30fps at 1920x1080 resolution, direct live streaming to monitor for viewing.
12. Should be provided with a 21 inch or above monitor for displaying the image.
13. Should provide dust cover for stand.

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail